

# ABX Minotrol Retic



**LOT** RX 443  
Rev 1

**CONTROL**

(Exp.) 2023-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.4	1.4	± 0.7	5.5	5.5	± 2.3	11.8	11.8	± 3.3			
RETIC	%	1.63	1.63	± 0.82	5.02	5.02	± 1.26	9.91	9.91	± 1.98			
RETIC	10 <sup>6</sup> /µl	0.056	0.056	± 0.028	0.170	0.170	± 0.043	0.332	0.332	± 0.066			
MFI MRI *	%	32.2	32.2	± 8.0	44.5	44.5	± 10.0	50.4	50.4	± 12.0			
GR RBC	10 <sup>6</sup> /µl	3.42	3.42	± 0.30	3.38	3.38	± 0.30	3.35	3.35	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.44 FRONT / RECTO 1300090305

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 443  
Rev 1

**CONTROL**

(Exp.) 2023-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA		XLR		PENTRA		XLR		PENTRA		XLR	
Méthode Manuelle / Manual RETIC	%	<b>1.4</b>			± 0.7	<b>5.5</b>			± 2.3	<b>11.8</b>			± 3.3
RETIC	%	<b>1.88</b>			± 0.94	<b>5.40</b>			± 1.35	<b>10.84</b>			± 2.17
RETIC	10 <sup>6</sup> /µl	<b>0.063</b>			± 0.032	<b>0.179</b>			± 0.045	<b>0.357</b>			± 0.071
MFI MRI *	%	<b>29.5</b>			± 8.0	<b>38.2</b>			± 10.0	<b>39.0</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.37</b>			± 0.30	<b>3.32</b>			± 0.30	<b>3.29</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.44 BACK / VERSO 1300090305

\* For research use only in Canada and USA