


**LOT** PX 443  
Rev 2


**CONTROL**

 (Exp.) 2023-11-05  
(YYYY - MM - DD)

		ABX Lysebio															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA					
		XL80	XLR				XL80	XLR				XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>3.1</b>	<b>3.1</b>			± 0.4	<b>8.6</b>	<b>8.6</b>			± 1.0	<b>18.8</b>	<b>18.8</b>			± 2.2	
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.24</b>	<b>2.24</b>			± 0.16	<b>4.63</b>	<b>4.63</b>			± 0.20	<b>5.20</b>	<b>5.20</b>			± 0.25	
HB HGB	g/dl	<b>5.8</b>	<b>5.8</b>			± 0.4	<b>13.7</b>	<b>13.7</b>			± 0.5	<b>16.9</b>	<b>16.9</b>			± 0.6	
	g/l	<b>58</b>	<b>58</b>			± 4	<b>137</b>	<b>137</b>			± 5	<b>169</b>	<b>169</b>			± 6	
	mmol/l	<b>3.60</b>	<b>3.60</b>			± 0.25	<b>8.51</b>	<b>8.51</b>			± 0.31	<b>10.49</b>	<b>10.49</b>			± 0.37	
HT HCT	%	<b>17.2</b>	<b>17.2</b>			± 1.5	<b>39.4</b>	<b>39.4</b>			± 2.0	<b>47.8</b>	<b>47.8</b>			± 2.5	
	l/l	<b>0.172</b>	<b>0.172</b>			± 0.015	<b>0.394</b>	<b>0.394</b>			± 0.020	<b>0.478</b>	<b>0.478</b>			± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	<b>77</b>	<b>77</b>			± 5	<b>85</b>	<b>85</b>			± 5	<b>92</b>	<b>92</b>			± 5	
TGMH MCH	pg	<b>25.9</b>	<b>25.9</b>			± 2.0	<b>29.6</b>	<b>29.6</b>			± 2.0	<b>32.5</b>	<b>32.5</b>			± 2.5	
	fmol	<b>1.61</b>	<b>1.61</b>			± 0.12	<b>1.84</b>	<b>1.84</b>			± 0.12	<b>2.02</b>	<b>2.02</b>			± 0.16	
CCMH MCHC	g/dl	<b>33.6</b>	<b>33.6</b>			± 3.0	<b>34.8</b>	<b>34.8</b>			± 3.0	<b>35.3</b>	<b>35.3</b>			± 3.0	
	g/l	<b>336</b>	<b>336</b>			± 30	<b>348</b>	<b>348</b>			± 30	<b>353</b>	<b>353</b>			± 30	
	mmol/l	<b>20.88</b>	<b>20.88</b>			± 1.86	<b>21.62</b>	<b>21.62</b>			± 1.86	<b>21.94</b>	<b>21.94</b>			± 1.86	
IDR RDW	%	<b>15.0</b>	<b>15.0</b>			± 4.0	<b>13.0</b>	<b>13.0</b>			± 4.0	<b>11.5</b>	<b>11.5</b>			± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>63</b>	<b>63</b>			± 20	<b>245</b>	<b>245</b>			± 30	<b>435</b>	<b>435</b>			± 50	
VPM MPV	µm <sup>3</sup> ; fl	<b>9.7</b>	<b>9.7</b>			± 2.0	<b>9.6</b>	<b>9.6</b>			± 2.0	<b>10.7</b>	<b>10.7</b>			± 2.0	
NEUT	#	<b>1.53</b>	<b>1.53</b>			± 0.35	<b>4.53</b>	<b>4.53</b>			± 0.90	<b>13.40</b>	<b>13.40</b>			± 1.90	
	%	<b>49.5</b>	<b>49.5</b>			± 10.0	<b>52.7</b>	<b>52.7</b>			± 10.0	<b>71.3</b>	<b>71.3</b>			± 10.0	
LYMPHO	#	<b>1.18</b>	<b>1.18</b>			± 0.33	<b>3.29</b>	<b>3.29</b>			± 0.70	<b>3.42</b>	<b>3.42</b>			± 1.50	
	%	<b>38.0</b>	<b>38.0</b>			± 12.0	<b>38.3</b>	<b>38.3</b>			± 8.0	<b>18.2</b>	<b>18.2</b>			± 8.0	
MONO	#	<b>0.13</b>	<b>0.13</b>			± 0.13	<b>0.26</b>	<b>0.26</b>			± 0.26	<b>0.45</b>	<b>0.45</b>			± 0.45	
	%	<b>4.3</b>	<b>4.3</b>			± 4.3	<b>3.0</b>	<b>3.0</b>			± 3.0	<b>2.4</b>	<b>2.4</b>			± 2.4	
EOS	#	<b>0.17</b>	<b>0.17</b>			± 0.17	<b>0.28</b>	<b>0.28</b>			± 0.28	<b>0.77</b>	<b>0.77</b>			± 0.77	
	%	<b>5.6</b>	<b>5.6</b>			± 5.6	<b>3.3</b>	<b>3.3</b>			± 3.3	<b>4.1</b>	<b>4.1</b>			± 4.1	
BASO	#	<b>0.08</b>	<b>0.08</b>			± 0.08	<b>0.23</b>	<b>0.23</b>			± 0.23	<b>0.75</b>	<b>0.75</b>			± 0.75	
	%	<b>2.6</b>	<b>2.6</b>			± 2.6	<b>2.7</b>	<b>2.7</b>			± 2.7	<b>4.0</b>	<b>4.0</b>			± 4.0	

**LOT** PX 443  
Rev 2

**CONTROL**

 (Exp.) 2023-11-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR		60 60C+ ES60	MS60	MSCR			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.0	3.0	3.0		± 0.4	8.8	8.5	8.5		± 1.0	19.2	18.7	18.7		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.29	2.26	2.26		± 0.16	4.64	4.62	4.57		± 0.20	5.20	5.21	5.14		± 0.25			
HB HGB	g/dl	5.8	5.8	5.8		± 0.4	13.6	13.5	13.6		± 0.5	16.8	16.7	16.7		± 0.6			
	g/l	58	58	58		± 4	136	135	136		± 5	168	167	167		± 6			
HT HCT	mmol/l	3.60	3.60	3.60		± 0.25	8.45	8.38	8.45		± 0.31	10.43	10.37	10.37		± 0.37			
	%	17.2	17.0	17.0		± 1.5	39.4	38.8	38.8		± 2.0	48.4	47.9	47.5		± 2.5			
VGM MCV	l/l	0.172	0.170	0.170		± 0.015	0.394	0.388	0.388		± 0.020	0.484	0.479	0.475		± 0.025			
	µm <sup>3</sup> ·fl	75	75	75.0		± 5	85	84	85.0		± 5	93	92	92.5		± 5			
TGMH MCH	pg	25.3	25.7	25.7		± 2.0	29.3	29.2	29.8		± 2.0	32.3	32.1	32.5		± 2.5			
	fmol	1.57	1.59	1.59		± 0.12	1.82	1.81	1.85		± 0.12	2.01	1.99	2.02		± 0.16			
CCMH MCHC	g/dl	33.8	34.2	34.2		± 3.0	34.5	34.8	35.0		± 3.0	34.7	34.8	35.1		± 3.0			
	g/l	338	342	342		± 30	345	348	350		± 30	347	348	351		± 30			
IDR RDW	mmol/l	20.97	21.25	21.25		± 1.86	21.41	21.60	21.74		± 1.86	21.57	21.64	21.81		± 1.86			
	%	16.0	16.0	14.5		± 4.0	12.5	12.5	12.0		± 4.0	11.5	12.0	11.0		± 4.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	69	64	65		± 20	251	247	233		± 30	467	465	430		± 50			
VPM MPV	µm <sup>3</sup> ·fl	9.4	9.3	8.5		± 2.0	8.5	8.8	8.3		± 2.0	9.7	9.8	9.1		± 2.0			
NEUT	#	1.50	1.45	1.50		± 0.35	4.58	4.37	4.38		± 0.90	13.56	13.26	13.37		± 1.90			
	%	49.9	48.2	50.0		± 10.0	52.0	51.4	51.5		± 10.0	70.6	70.9	71.5		± 10.0			
LYMPHO	#	1.13	1.17	1.16		± 0.33	3.40	3.34	3.36		± 0.70	3.53	3.48	3.46		± 1.50			
	%	37.7	38.9	38.5		± 12.0	38.6	39.3	39.5		± 8.0	18.4	18.6	18.5		± 8.0			
MONO	#	0.13	0.13	0.12		± 0.12	0.26	0.26	0.26		± 0.26	0.44	0.43	0.37		± 0.37			
	%	4.3	4.3	4.0		± 4.0	3.0	3.0	3.0		± 3.0	2.3	2.3	2.0		± 2.0			
EOS	#	0.17	0.18	0.14		± 0.14	0.32	0.30	0.26		± 0.26	0.88	0.79	0.75		± 0.75			
	%	5.5	6.0	4.5		± 4.5	3.6	3.5	3.0		± 3.0	4.6	4.2	4.0		± 4.0			
BASO	#	0.08	0.08	0.09		± 0.08	0.25	0.24	0.26		± 0.24	0.79	0.75	0.75		± 0.75			
	%	2.6	2.6	3.0		± 2.6	2.8	2.8	3.0		± 2.8	4.1	4.0	4.0		± 4.0			

Ref: TEMP-0821 Rev.49 FRONT / RECTO