


**LOT** PX 443

**CONTROL**

 (Exp.) 2023-11-05  
(YYYY-MM-DD)

Rev 2

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.0	3.0			± 0.4	8.8	8.8			± 1.0	19.3	19.3			± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.30	2.30			± 0.16	4.70	4.70			± 0.20	5.38	5.38			± 0.25			
HB HGB	g/dl	5.9	5.9			± 0.4	13.6	13.6			± 0.5	16.7	16.7			± 0.6			
	g/l	59	59			± 4	136	136			± 5	167	167			± 6			
	mmol/l	3.66	3.66			± 0.25	8.45	8.45			± 0.31	10.37	10.37			± 0.37			
HT HCT	%	17.9	17.9			± 1.5	40.0	40.0			± 2.0	49.0	49.0			± 2.5			
	l/l	0.179	0.179			± 0.015	0.400	0.400			± 0.020	0.490	0.490			± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	78	78			± 5	85	85			± 5	91	91			± 5			
TGMH MCH	pg	25.7	25.7			± 2.0	28.9	28.9			± 2.0	31.0	31.0			± 2.5			
	fmol	1.59	1.59			± 0.12	1.80	1.80			± 0.12	1.93	1.93			± 0.16			
CCMH MCHC	g/dl	32.9	32.9			± 3.0	34.0	34.0			± 3.0	34.1	34.1			± 3.0			
	g/l	329	329			± 30	340	340			± 30	341	341			± 30			
	mmol/l	20.42	20.42			± 1.86	21.14	21.14			± 1.86	21.18	21.18			± 1.86			
IDR RDW	%	18.0	18.0			± 4.0	15.0	15.0			± 4.0	13.0	13.0			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	69	69			± 20	255	255			± 30	475	475			± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.8	9.8			± 2.0	9.2	9.2			± 2.0	10.3	10.3			± 2.0			
NEUT	#	1.64	1.64			± 0.35	4.98	4.98			± 0.90	14.00	14.00			± 1.90			
	%	54.8	54.8			± 10.0	56.6	56.6			± 10.0	72.6	72.6			± 10.0			
LYMPHO	#	0.86	0.86			± 0.33	2.84	2.84			± 0.70	2.99	2.99			± 1.50			
	%	28.7	28.7			± 12.0	32.3	32.3			± 8.0	15.5	15.5			± 8.0			
MONO	#	0.17	0.17			± 0.17	0.34	0.34			± 0.34	0.56	0.56			± 0.56			
	%	5.6	5.6			± 5.6	3.9	3.9			± 3.9	2.9	2.9			± 2.9			
EOS	#	0.19	0.19			± 0.19	0.29	0.29			± 0.29	0.85	0.85			± 0.85			
	%	6.2	6.2			± 6.2	3.3	3.3			± 3.3	4.4	4.4			± 4.4			
BASO	#	0.14	0.14			± 0.14	0.34	0.34			± 0.34	0.89	0.89			± 0.89			
	%	4.7	4.7			± 4.7	3.9	3.9			± 3.9	4.6	4.6			± 4.6			

Ref: TEMP-0821 Rev.49 BACK / VERSO