


LOT PX 443
Rev 2


CONTROL

 (Exp.) 2023-11-05
(YYYY - MM - DD)

		ABX Lysebio															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA					
		XL80	XLR				XL80	XLR				XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.1	3.1			± 0.4	8.6	8.6			± 1.0	18.8	18.8			± 2.2	
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.24	2.24			± 0.16	4.63	4.63			± 0.20	5.20	5.20			± 0.25	
HB HGB	g/dl	5.8	5.8			± 0.4	13.7	13.7			± 0.5	16.9	16.9			± 0.6	
	g/l	58	58			± 4	137	137			± 5	169	169			± 6	
	mmol/l	3.60	3.60			± 0.25	8.51	8.51			± 0.31	10.49	10.49			± 0.37	
HT HCT	%	17.2	17.2			± 1.5	39.4	39.4			± 2.0	47.8	47.8			± 2.5	
	l/l	0.172	0.172			± 0.015	0.394	0.394			± 0.020	0.478	0.478			± 0.025	
VGM MCV	µm ³ fl	77	77			± 5	85	85			± 5	92	92			± 5	
TGMH MCH	pg	25.9	25.9			± 2.0	29.6	29.6			± 2.0	32.5	32.5			± 2.5	
	fmol	1.61	1.61			± 0.12	1.84	1.84			± 0.12	2.02	2.02			± 0.16	
CCMH MCHC	g/dl	33.6	33.6			± 3.0	34.8	34.8			± 3.0	35.3	35.3			± 3.0	
	g/l	336	336			± 30	348	348			± 30	353	353			± 30	
	mmol/l	20.88	20.88			± 1.86	21.62	21.62			± 1.86	21.94	21.94			± 1.86	
IDR RDW	%	15.0	15.0			± 4.0	13.0	13.0			± 4.0	11.5	11.5			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	63	63			± 20	245	245			± 30	435	435			± 50	
VPM MPV	µm ³ fl	9.7	9.7			± 2.0	9.6	9.6			± 2.0	10.7	10.7			± 2.0	
NEUT	#	1.53	1.53			± 0.35	4.53	4.53			± 0.90	13.40	13.40			± 1.90	
	%	49.5	49.5			± 10.0	52.7	52.7			± 10.0	71.3	71.3			± 10.0	
LYMPHO	#	1.18	1.18			± 0.33	3.29	3.29			± 0.70	3.42	3.42			± 1.50	
	%	38.0	38.0			± 12.0	38.3	38.3			± 8.0	18.2	18.2			± 8.0	
MONO	#	0.13	0.13			± 0.13	0.26	0.26			± 0.26	0.45	0.45			± 0.45	
	%	4.3	4.3			± 4.3	3.0	3.0			± 3.0	2.4	2.4			± 2.4	
EOS	#	0.17	0.17			± 0.17	0.28	0.28			± 0.28	0.77	0.77			± 0.77	
	%	5.6	5.6			± 5.6	3.3	3.3			± 3.3	4.1	4.1			± 4.1	
BASO	#	0.08	0.08			± 0.08	0.23	0.23			± 0.23	0.75	0.75			± 0.75	
	%	2.6	2.6			± 2.6	2.7	2.7			± 2.7	4.0	4.0			± 4.0	

LOT PX 443
Rev 2

CONTROL

 (Exp.) 2023-11-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0	3.0		± 0.4	8.8	8.5	8.5		± 1.0	19.2	18.7	18.7		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.29	2.26	2.26		± 0.16	4.64	4.62	4.57		± 0.20	5.20	5.21	5.14		± 0.25			
HB HGB	g/dl	5.8	5.8	5.8		± 0.4	13.6	13.5	13.6		± 0.5	16.8	16.7	16.7		± 0.6			
	g/l	58	58	58		± 4	136	135	136		± 5	168	167	167		± 6			
HT HCT	mmol/l	3.60	3.60	3.60		± 0.25	8.45	8.38	8.45		± 0.31	10.43	10.37	10.37		± 0.37			
	%	17.2	17.0	17.0		± 1.5	39.4	38.8	38.8		± 2.0	48.4	47.9	47.5		± 2.5			
VGM MCV	l/l	0.172	0.170	0.170		± 0.015	0.394	0.388	0.388		± 0.020	0.484	0.479	0.475		± 0.025			
	µm ³ ·fl	75	75	75.0		± 5	85	84	85.0		± 5	93	92	92.5		± 5			
TGMH MCH	pg	25.3	25.7	25.7		± 2.0	29.3	29.2	29.8		± 2.0	32.3	32.1	32.5		± 2.5			
	fmol	1.57	1.59	1.59		± 0.12	1.82	1.81	1.85		± 0.12	2.01	1.99	2.02		± 0.16			
CCMH MCHC	g/dl	33.8	34.2	34.2		± 3.0	34.5	34.8	35.0		± 3.0	34.7	34.8	35.1		± 3.0			
	g/l	338	342	342		± 30	345	348	350		± 30	347	348	351		± 30			
IDR RDW	mmol/l	20.97	21.25	21.25		± 1.86	21.41	21.60	21.74		± 1.86	21.57	21.64	21.81		± 1.86			
	%	16.0	16.0	14.5		± 4.0	12.5	12.5	12.0		± 4.0	11.5	12.0	11.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	64	65		± 20	251	247	233		± 30	467	465	430		± 50			
VPM MPV	µm ³ ·fl	9.4	9.3	8.5		± 2.0	8.5	8.8	8.3		± 2.0	9.7	9.8	9.1		± 2.0			
NEUT	#	1.50	1.45	1.50		± 0.35	4.58	4.37	4.38		± 0.90	13.56	13.26	13.37		± 1.90			
	%	49.9	48.2	50.0		± 10.0	52.0	51.4	51.5		± 10.0	70.6	70.9	71.5		± 10.0			
LYMPHO	#	1.13	1.17	1.16		± 0.33	3.40	3.34	3.36		± 0.70	3.53	3.48	3.46		± 1.50			
	%	37.7	38.9	38.5		± 12.0	38.6	39.3	39.5		± 8.0	18.4	18.6	18.5		± 8.0			
MONO	#	0.13	0.13	0.12		± 0.12	0.26	0.26	0.26		± 0.26	0.44	0.43	0.37		± 0.37			
	%	4.3	4.3	4.0		± 4.0	3.0	3.0	3.0		± 3.0	2.3	2.3	2.0		± 2.0			
EOS	#	0.17	0.18	0.14		± 0.14	0.32	0.30	0.26		± 0.26	0.88	0.79	0.75		± 0.75			
	%	5.5	6.0	4.5		± 4.5	3.6	3.5	3.0		± 3.0	4.6	4.2	4.0		± 4.0			
BASO	#	0.08	0.08	0.09		± 0.08	0.25	0.24	0.26		± 0.24	0.79	0.75	0.75		± 0.75			
	%	2.6	2.6	3.0		± 2.6	2.8	2.8	3.0		± 2.8	4.1	4.0	4.0		± 4.0			

Ref: TEMP-0821 Rev.49 FRONT / RECTO