



LOT MX 443 L
 Rev 1

CONTROL L



2023-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.0			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.32			2.29					± 0.17
HB HGB	g/dl	6.1	6.1			6.1					± 0.5
	g/l	61	61			61					± 5
	mmol/l	3.79	3.79			3.79					± 0.31
HT HCT	%	16.8	16.5			16.3					± 2.0
	l/l	0.168	0.165			0.163					± 0.020
VGM MCV	µm ³ ; fl	72	71			71					± 4
TGMH MCH	pg	26.2	26.3			26.6					± 2.0
	fmol	1.63	1.63			1.65					± 0.12
CCMH MCHC	g/dl	36.4	37.0			37.5					± 3.0
	g/l	364	370			375					± 30
	mmol/l	22.58	23.00			23.30					± 1.86
IDR RDW	%	15.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	77	75			76					± 22
VPM MPV	µm ³ ; fl	8.2	8.0			7.5					± 2.0
LYMPHO	#	1.33	1.28			1.35					± 0.40
	%	64.5	65.0			65.5					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.5	9.0			8.5					± 6.0
GRANULO	#	0.57	0.52			0.55					± 0.40
	%	27.0	26.0			26.0					± 7.0

LOT MX 443 N
 Rev 1

CONTROL N

(Exp.)

2023-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.6	7.3			7.5					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.52	4.48			4.49					± 0.26
HB HGB	g/dl	13.5	13.4			13.5					± 0.7
	g/l	135	134			135					± 7
	mmol/l	8.38	8.32			8.38					± 0.43
HT HCT	%	37.1	36.7			36.8					± 2.7
	l/l	0.371	0.367			0.368					± 0.027
VGM MCV	µm ³ ; fl	82	82			82					± 4
TGMH MCH	pg	29.9	29.9			30.1					± 2.0
	fmol	1.85	1.86			1.87					± 0.12
CCMH MCHC	g/dl	36.4	36.5			36.7					± 3.0
	g/l	364	365			367					± 30
	mmol/l	22.62	22.65			22.77					± 1.86
IDR RDW	%	14.5	13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	251	251			246					± 40
VPM MPV	µm ³ ; fl	8.0	7.5			7.5					± 2.0
LYMPHO	#	2.51	2.41			2.48					± 0.40
	%	33.0	33.0			33.0					± 5.0
MONO	#	0.57	0.55			0.53					± 0.30
	%	7.5	7.5			7.0					± 4.0
GRANULO	#	4.52	4.34			4.50					± 0.60
	%	59.5	59.5			60.0					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104144



LOT

MX 443 H

CONTROL H

(Exp.)

2023-12-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.1	20.2			20.2					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.70	5.66			5.69					± 0.32
HB HGB	g/dl	18.0	18.0			18.0					± 0.9
	g/l	180	180			180					± 9
	mmol/l	11.18	11.18			11.18					± 0.56
HT HCT	%	50.2	49.8			50.6					± 3.4
	l/l	0.502	0.498			0.506					± 0.034
VGM MCV	µm ³ ; fl	88	88			89					± 4
TGMH MCH	pg	31.6	31.8			31.6					± 2.0
	fmol	1.96	1.97			1.96					± 0.12
CCMH MCHC	g/dl	35.9	36.1			35.5					± 3.0
	g/l	359	361			355					± 30
	mmol/l	22.28	22.44			22.07					± 1.86
IDR RDW	%	14.0	13.0			12.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	492	481			488					± 65
VPM MPV	µm ³ ; fl	7.7	7.5			7.0					± 2.0
LYMPHO	#	3.52	3.43			3.33					± 1.00
	%	17.5	17.0			16.5					± 5.0
MONO	#	1.01	1.21			1.01					± 0.80
	%	5.0	6.0			5.0					± 4.0
GRANULO	#	15.58	15.55			15.86					± 1.40
	%	77.5	77.0			78.5					± 7.0