





LOT		MX 443 L		Rev 1		CONTROL		L		 (Exp.)		2023-12-05 (YYYY - MM - DD)	
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE		
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA			
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60							
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	2.1	2.0			2.1						± 0.4	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.32			2.29						± 0.17	
HB HGB	g/dl	6.1	6.1			6.1						± 0.5	
	g/l	61	61			61						± 5	
	mmol/l	3.79	3.79			3.79						± 0.31	
HT HCT	%	16.8	16.5			16.3						± 2.0	
	l/l	0.168	0.165			0.163						± 0.020	
VGM MCV	µm ³ ; fl	72	71			71						± 4	
TGMH MCH	pg	26.2	26.3			26.6						± 2.0	
	fmol	1.63	1.63			1.65						± 0.12	
CCMH MCHC	g/dl	36.4	37.0			37.5						± 3.0	
	g/l	364	370			375						± 30	
	mmol/l	22.58	23.00			23.30						± 1.86	
IDR RDW	%	15.0	14.0			14.0						± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	77	75			76						± 22	
VPM MPV	µm ³ ; fl	8.2	8.0			7.5						± 2.0	
LYMPHO	#	1.33	1.28			1.35						± 0.40	
	%	64.5	65.0			65.5						± 8.0	
MONO	#	0.20	0.20			0.20						± 0.20	
	%	8.5	9.0			8.5						± 6.0	
GRANULO	#	0.57	0.52			0.55						± 0.40	
	%	27.0	26.0			26.0						± 7.0	

LOT		MX 443 N		CONTROL		N		 (Exp.)		2023-12-05 (YYYY-MM-DD)											
PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse								TOLERANCES TOLERANCE									
				MICROS		MICROS		MICROS		MICROS				ARGOS - HELIOS		PENTRA		PENTRA		PENTRA	
				ES60 Care ST		CRP 200				60 8P - 60 18P ADVIA 60											
GB	WBC	10 ³ /mm ³ ; 10 ⁹ /l	7.6	7.3			7.5											± 0.8			
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.52	4.48			4.49											± 0.26			
HB	HGB	g/dl	13.5	13.4			13.5											± 0.7			
		g/l	135	134			135										± 7				
		mmol/l	8.38	8.32			8.38										± 0.43				
HT	HCT	%	37.1	36.7			36.8										± 2.7				
		l/l	0.371	0.367			0.368										± 0.027				
VGM	MCV	µm ³ ; fl	82	82			82											± 4			
TGMH	MCH	pg	29.9	29.9			30.1											± 2.0			
		fmol	1.85	1.86			1.87											± 0.12			
CCMH	MCHC	g/dl	36.4	36.5			36.7											± 3.0			
		g/l	364	365			367										± 30				
		mmol/l	22.62	22.65			22.77										± 1.86				
IDR	RDW	%	14.5	13.5			13.5										± 3.0				
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	251	251			246											± 40			
VPM	MPV	µm ³ ; fl	8.0	7.5			7.5											± 2.0			
LYMPHO	#	2.51	2.41			2.48												± 0.40			
	%	33.0	33.0			33.0												± 5.0			
MONO	#	0.57	0.55			0.53												± 0.30			
	%	7.5	7.5			7.0												± 4.0			
GRANULO	#	4.52	4.34			4.50												± 0.60			
	%	59.5	59.5			60.0												± 6.0			

Rel: TEMP-0830 Rev.42 FRONT / RECTO 1300104144



LOT		MX 443 H		Rev 1		CONTROL		H		<div><div></div>(Exp.)</div>		2023-12-05 (YYYY - MM - DD)	
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE		
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA			
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60							
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	20.1	20.2			20.2						± 1.7	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.70	5.66			5.69						± 0.32	
HB HGB	g/dl	18.0	18.0			18.0						± 0.9	
	g/l	180	180			180						± 9	
	mmol/l	11.18	11.18			11.18						± 0.56	
HT HCT	%	50.2	49.8			50.6						± 3.4	
	l/l	0.502	0.498			0.506						± 0.034	
VGM MCV	µm ³ ; fl	88	88			89						± 4	
TGMH MCH	pg	31.6	31.8			31.6						± 2.0	
	fmol	1.96	1.97			1.96						± 0.12	
CCMH MCHC	g/dl	35.9	36.1			35.5						± 3.0	
	g/l	359	361			355						± 30	
	mmol/l	22.28	22.44			22.07						± 1.86	
IDR RDW	%	14.0	13.0			12.5						± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	492	481			488						± 65	
VPM MPV	µm ³ ; fl	7.7	7.5			7.0						± 2.0	
LYMPHO	#	3.52	3.43			3.33						± 1.00	
	%	17.5	17.0			16.5						± 5.0	
MONO	#	1.01	1.21			1.01						± 0.80	
	%	5.0	6.0			5.0						± 4.0	
GRANULO	#	15.58	15.55			15.86						± 1.40	
	%	77.5	77.0			78.5						± 7.0	