

LOT MX 443 L
 Rev 1

CONTROL L

(Exp.)

2023-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.34	2.32			2.31					± 0.17
HB HGB	g/dl	6.0	6.0			5.7					± 0.5
	g/l	60	60			57					± 5
	mmol/l	3.73	3.71			3.52					± 0.31
HT HCT	%	16.6	16.5			16.4					± 2.0
	l/l	0.166	0.165			0.164					± 0.020
VGM MCV	µm ³ ; fl	71	71			71					± 4
TGMH MCH	pg	25.6	25.8			24.6					± 2.0
	fmol	1.59	1.60			1.53					± 0.12
CCMH MCHC	g/dl	36.1	36.3			34.6					± 3.0
	g/l	361	363			346					± 30
	mmol/l	22.43	22.54			21.48					± 1.86
IDR RDW	%	15.0	14.0			15.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	75			76					± 22
VPM MPV	µm ³ ; fl	8.3	8.0			7.5					± 2.0
LYMPHO	#	1.28	1.29			1.35					± 0.40
	%	67.0	65.7			67.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	7.0	9.0			7.0					± 6.0
GRANULO	#	0.52	0.51			0.55					± 0.40
	%	26.0	25.5			26.0					± 7.0

LOT MX 443 N
 Rev 1

CONTROL N

(Exp.)

2023-12-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5	7.3			7.3					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.57	4.48			4.52					± 0.26
HB HGB	g/dl	13.5	13.4			13.2					± 0.7
	g/l	135	134			132					± 7
	mmol/l	8.38	8.32			8.22					± 0.43
HT HCT	%	37.5	36.7			37.5					± 2.7
	l/l	0.375	0.367			0.375					± 0.027
VGM MCV	µm ³ ; fl	82	82			83					± 4
TGMH MCH	pg	29.5	29.9			29.3					± 2.0
	fmol	1.83	1.86			1.82					± 0.12
CCMH MCHC	g/dl	36.0	36.5			35.3					± 3.0
	g/l	360	365			353					± 30
	mmol/l	22.37	22.65			21.90					± 1.86
IDR RDW	%	14.0	13.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	243	251			246					± 40
VPM MPV	µm ³ ; fl	8.2	7.5			7.5					± 2.0
LYMPHO	#	2.49	2.55			2.48					± 0.40
	%	34.5	35.0			34.0					± 5.0
MONO	#	0.45	0.49			0.44					± 0.30
	%	6.0	6.7			6.0					± 4.0
GRANULO	#	4.46	4.26			4.38					± 0.60
	%	59.5	58.3			60.0					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143



LOT		MX 443 H	CONTROL H				(Exp.)				2023-12-05 (YYYY-MM-DD)
Rev 1		Minilysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.2	20.2			19.6					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.75	5.66			5.68					± 0.32
HB HGB	g/dl	18.5	18.0			17.8					± 0.9
	g/l	185	180			178					± 9
	mmol/l	11.49	11.18			11.07					± 0.56
HT HCT	%	51.2	49.8			50.6					± 3.4
	l/l	0.512	0.498			0.506					± 0.034
VGM MCV	µm ³ ; fl	89	88			89					± 4
TGMH MCH	pg	32.2	31.8			31.4					± 2.0
	fmol	2.00	1.97			1.95					± 0.12
CCMH MCHC	g/dl	36.2	36.1			35.3					± 3.0
	g/l	362	361			353					± 30
	mmol/l	22.45	22.44			21.89					± 1.86
IDR RDW	%	13.5	13.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	469	481			488					± 65
VPM MPV	µm ³ ; fl	8.0	7.5			7.0					± 2.0
LYMPHO	#	3.64	3.64			3.53					± 1.00
	%	18.0	18.0			18.0					± 5.0
MONO	#	1.01	1.31			0.88					± 0.80
	%	5.0	6.5			4.5					± 4.0
GRANULO	#	15.55	15.24			15.19					± 1.40
	%	77.0	75.5			77.5					± 7.0