



LOT		MX 443	L	CONTROL		L	 (Exp.)				2023-12-05 (YYYY - MM - DD)	
		Rev 1		LyseVet								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
		Scil Vet ABC	Scil Vet ABC Plus ESV 60									
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	1.7	1.9								± 0.4	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.34	2.33								± 0.15	
HB HGB	g/dl	6.0	6.0								± 0.4	
	g/l	60	60								± 4	
	mmol/l	3.73	3.73								± 0.25	
HT HCT	%	16.6	16.5								± 2.0	
	l/l	0.166	0.165								± 0.020	
VGM MCV	µm ³ ; fl	71	71								± 4	
TGMH MCH	pg	25.6	25.8								± 2.0	
	fmol	1.59	1.60								± 0.12	
CCMH MCHC	g/dl	36.1	36.3								± 3.0	
	g/l	361	363								± 30	
	mmol/l	22.43	22.52								± 1.86	
IDR RDW	%	N/A	15.0								± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	69	77								± 20	
VPM MPV	µm ³ ; fl	N/A	8.4								± 2.0	
LYMPHO	#	N/A	1.10								± 0.40	
	%	N/A	59.0								± 8.0	
MONO	#	N/A	0.20								± 0.20	
	%	N/A	9.5								± 6.0	
GRANULO	#	N/A	0.60								± 0.40	
	%	N/A	31.5								± 7.0	

LOT		MX 443	N	CONTROL		N	 (Exp.)				2023-12-05 (YYYY-MM-DD)
Rev 1		LyseVet									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE
		Scil Vet ABC	Scil Vet ABC Plus ESV 60								
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	6.8	7.1								± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.51	4.52								± 0.18
HB HGB	g/dl	13.2	13.3								± 0.5
	g/l	132	133								± 5
	mmol/l	8.20	8.26								± 0.31
HT HCT	%	36.5	37.1								± 2.5
	l/l	0.365	0.371								± 0.025
VGM MCV	µm ³ ; fl	81	82								± 4
TGMH MCH	pg	29.3	29.4								± 2.0
	fmol	1.82	1.83								± 0.12
CCMH MCHC	g/dl	36.1	35.9								± 3.0
	g/l	361	359								± 30
	mmol/l	22.44	22.28								± 1.86
IDR RDW	%	N/A	14.0								± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	231	249								± 40
VPM MPV	µm ³ ; fl	N/A	8.5								± 2.0
LYMPHO	#	N/A	1.88								± 0.40
	%	N/A	26.5								± 5.0
MONO	#	N/A	0.57								± 0.30
	%	N/A	8.0								± 4.0
GRANULO	#	N/A	4.65								± 0.60
	%	N/A	65.5								± 6.0

Ref: TEMP-0830 Rev.42 BACK / VERSO 1300104143

LOT

MX 443 H
Rev 1

CONTROL

H

 (Exp.)

2023-12-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	19.6	19.5							± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.72	5.70							± 0.20
HB HGB	g/dl	18.1	18.0							± 0.6
	g/l	181	180							± 6
	mmol/l	11.24	11.18							± 0.37
HT HCT	%	49.8	50.7							± 3.0
	l/l	0.498	0.507							± 0.030
VGM MCV	µm ³ ; fl	87	89							± 4
TGMH MCH	pg	31.6	31.6							± 2.0
	fmol	1.97	1.96							± 0.12
CCMH MCHC	g/dl	36.4	35.5							± 3.0
	g/l	364	355							± 30
	mmol/l	22.59	22.03							± 1.86
IDR RDW	%	N/A	13.0							± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	467	491							± 45
VPM MPV	µm ³ ; fl	N/A	8.0							± 2.0
LYMPHO	#	N/A	2.34							± 1.00
	%	N/A	12.0							± 5.0
MONO	#	N/A	0.98							± 0.80
	%	N/A	5.0							± 4.0
GRANULO	#	N/A	16.19							± 1.40
	%	N/A	83.0							± 7.0