



LOT		MX 443 L		CONTROL		L		 (Exp.)		2023-12-05 (YYYY-MM-DD)							
Rev 1		ABX Lysebio															
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE							
		EMI CRP															
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l									2.2							± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l									2.34							± 0.15
HB HGB	g/dl									6.1							± 0.4
	g/l									61							± 4
	mmol/l									3.79							± 0.25
HT HCT	%									16.4							± 2.0
	l/l									0.164							± 0.020
VGM MCV	µm ³ ; fl									70							± 4
TGMH MCH	pg									26.1							± 2.0
	fmol									1.62							± 0.12
CCMH MCHC	g/dl									37.2							± 3.0
	g/l									372							± 30
	mmol/l									23.13							± 1.86
IDR RDW	%									14.0							± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l									71							± 20
VPM MPV	µm ³ ; fl									7.5							± 2.0
LYMPHO	#									1.41							± 0.40
	%									65.0							± 8.0
MONO	#									0.20							± 0.20
	%									7.5							± 6.0
GRANULO	#									0.60							± 0.40
	%									27.5							± 7.0

LOT		MX 443	N	CONTROL		N				 (Exp.)	2023-12-05 (YYYY-MM-DD)	
		Rev 1			ABX Lysebio							
PARAMETRES PARAMETERS	UNITES UNITS	MICROS									TOLERANCES TOLERANCE	
				EMI CRP								
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			7.8							± 0.8	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			4.53							± 0.18	
HB HGB	g/dl			13.4							± 0.5	
	g/l			134							± 5	
	mmol/l			8.32							± 0.31	
HT HCT	%			36.7							± 2.5	
	l/l			0.367							± 0.025	
VGM MCV	µm ³ ; fl			81							± 4	
TGMH MCH	pg			29.6							± 2.0	
	fmol			1.84							± 0.12	
CCMH MCHC	g/dl			36.5							± 3.0	
	g/l			365							± 30	
	mmol/l			22.68							± 1.86	
IDR RDW	%			13.0							± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			241							± 40	
VPM MPV	µm ³ ; fl			8.1							± 2.0	
LYMPHO	#			2.65							± 0.40	
	%			34.0							± 5.0	
MONO	#			0.51							± 0.30	
	%			6.5							± 4.0	
GRANULO	#			4.64							± 0.60	
	%			59.5							± 6.0	

Ret: TEMP-0830 Rev.42 BACK / VERSO 1300104144

LOT MX 443 H
Rev 1

CONTROL H

 (Exp.)

2023-12-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			20.3						± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			5.60						± 0.20
HB HGB	g/dl			17.8						± 0.6
	g/l			178						± 6
	mmol/l			11.05						± 0.37
HT HCT	%			49.3						± 3.0
	l/l			0.493						± 0.030
VGM MCV	µm ³ ; fl			88						± 4
TGMH MCH	pg			31.8						± 2.0
	fmol			1.97						± 0.12
CCMH MCHC	g/dl			36.1						± 3.0
	g/l			361						± 30
	mmol/l			22.43						± 1.86
IDR RDW	%			12.5						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			476						± 45
VPM MPV	µm ³ ; fl			8.0						± 2.0
LYMPHO	#			3.76						± 1.00
	%			18.5						± 5.0
MONO	#			1.02						± 0.80
	%			5.0						± 4.0
GRANULO	#			15.53						± 1.40
	%			76.5						± 7.0