

ABX Minotrol CRP



LOT

MC 429

CONTROL



(Exp.) 2024-01-05
(YYYY - MM - DD)

Rev 2

		ABX Lyse																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
		MICROS						MICROS						MICROS					
		CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.4	7.6			± 1.0	19.9	20.4			± 2.2			
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.34	2.30			± 0.15	4.65	4.65			± 0.18	5.82	5.84			± 0.25			
HB HGB	g/dl	5.9	5.9			± 0.4	13.7	13.7			± 0.5	18.1	18.1			± 0.7			
	g/l	59	59			± 4	137	137			± 5	181	181			± 7			
	mmol/l	3.66	3.66			± 0.25	8.51	8.51			± 0.31	11.24	11.24			± 0.44			
HT HCT	%	16.4	16.1			± 2.0	37.7	37.7			± 2.5	51.2	50.8			± 3.0			
	l/l	0.164	0.161			± 0.020	0.377	0.377			± 0.025	0.512	0.508			± 0.030			
VGM MCV	µm ³ ; fl	70	70			± 4	81	81			± 4	88	87			± 4			
TGMH MCH	pg	25.2	25.7			± 2.4	29.5	29.5			± 2.4	31.1	31.0			± 2.8			
	fmol	1.57	1.59			± 0.15	1.83	1.83			± 0.15	1.93	1.92			± 0.18			
CCMH MCHC	g/dl	36.0	36.6			± 3.0	36.4	36.4			± 3.0	35.3	35.6			± 3.0			
	g/l	360	366			± 30	364	364			± 30	353	356			± 30			
	mmol/l	22.37	22.76			± 1.86	22.59	22.59			± 1.86	21.95	22.12			± 1.86			
IDR RDW	%	14.0	14.5			± 3.0	13.5	13.5			± 3.0	13.0	12.5			± 3.0			
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	83	83			± 20	262	263			± 40	510	519			± 60			
VPM MPV	µm ³ ; fl	8.8	8.0			± 2.0	7.9	7.5			± 2.0	7.7	7.1			± 2.0			
LYMPHO	#	1.4	1.3			± 0.2	2.6	2.6			± 0.5	3.1	3.1			± 1.1			
	%	67.5	66.0			± 8.0	35.0	34.5			± 6.0	15.5	15.0			± 5.0			
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.5	1.1			± 1.0			
	%	9.0	9.0			± 7.0	7.5	7.0			± 5.0	7.5	5.5			± 5.0			
GRANULO	#	0.5	0.5			± 0.2	4.3	4.4			± 0.5	15.3	16.2			± 1.5			
	%	23.5	25.0			± 7.0	57.5	58.0			± 6.0	77.0	79.5			± 7.0			
CRP	mg/dl	0.42	N/A			± 0.30	2.11	N/A			± 0.70	3.97	N/A			± 1.00			
	mg/l	4.2	N/A			± 3.0	21.1	N/A			± 7.0	39.7	N/A			± 10.0			

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170