

ABX Minotrol CRP



LOT

MC 429

CONTROL



(Exp.) 2024-01-05
(YYYY - MM - DD)

Rev 2

		ABX Lyse															
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0			± 0.4	7.4	7.6			± 1.0	19.9	20.4			± 2.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.34	2.30			± 0.15	4.65	4.65			± 0.18	5.82	5.84			± 0.25	
HB HGB	g/dl	5.9	5.9			± 0.4	13.7	13.7			± 0.5	18.1	18.1			± 0.7	
	g/l	59	59			± 4	137	137			± 5	181	181			± 7	
	mmol/l	3.66	3.66			± 0.25	8.51	8.51			± 0.31	11.24	11.24			± 0.44	
HT HCT	%	16.4	16.1			± 2.0	37.7	37.7			± 2.5	51.2	50.8			± 3.0	
	l/l	0.164	0.161			± 0.020	0.377	0.377			± 0.025	0.512	0.508			± 0.030	
VGM MCV	µm ³ ; fl	70	70			± 4	81	81			± 4	88	87			± 4	
TGMH MCH	pg	25.2	25.7			± 2.4	29.5	29.5			± 2.4	31.1	31.0			± 2.8	
	fmol	1.57	1.59			± 0.15	1.83	1.83			± 0.15	1.93	1.92			± 0.18	
CCMH MCHC	g/dl	36.0	36.6			± 3.0	36.4	36.4			± 3.0	35.3	35.6			± 3.0	
	g/l	360	366			± 30	364	364			± 30	353	356			± 30	
	mmol/l	22.37	22.76			± 1.86	22.59	22.59			± 1.86	21.95	22.12			± 1.86	
IDR RDW	%	14.0	14.5			± 3.0	13.5	13.5			± 3.0	13.0	12.5			± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	83	83			± 20	262	263			± 40	510	519			± 60	
VPM MPV	µm ³ ; fl	8.8	8.0			± 2.0	7.9	7.5			± 2.0	7.7	7.1			± 2.0	
LYMPHO	#	1.4	1.3			± 0.2	2.6	2.6			± 0.5	3.1	3.1			± 1.1	
	%	67.5	66.0			± 8.0	35.0	34.5			± 6.0	15.5	15.0			± 5.0	
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.5	1.1			± 1.0	
	%	9.0	9.0			± 7.0	7.5	7.0			± 5.0	7.5	5.5			± 5.0	
GRANULO	#	0.5	0.5			± 0.2	4.3	4.4			± 0.5	15.3	16.2			± 1.5	
	%	23.5	25.0			± 7.0	57.5	58.0			± 6.0	77.0	79.5			± 7.0	
CRP	mg/dl	0.42	N/A			± 0.30	2.11	N/A			± 0.70	3.97	N/A			± 1.00	
	mg/l	4.2	N/A			± 3.0	21.1	N/A			± 7.0	39.7	N/A			± 10.0	

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170