

# ABX Minotrol CRP



**LOT**

**MC 429**

**CONTROL**



**(Exp.) 2024-01-05**  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.0</b>		<b>2.1</b>		± 0.4	<b>7.6</b>	<b>7.6</b>		<b>7.9</b>		± 1.0	<b>20.2</b>	<b>20.7</b>		<b>21.1</b>		± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.35</b>	<b>2.34</b>		<b>2.31</b>		± 0.15	<b>4.61</b>	<b>4.58</b>		<b>4.61</b>		± 0.18	<b>5.68</b>	<b>5.73</b>		<b>5.76</b>		± 0.25
HB	HGB	g/dl	<b>6.1</b>	<b>6.0</b>		<b>5.8</b>		± 0.4	<b>13.5</b>	<b>13.5</b>		<b>13.7</b>		± 0.5	<b>17.8</b>	<b>17.9</b>		<b>18.2</b>		± 0.7
		g/l	<b>61</b>	<b>60</b>		<b>58</b>		± 4	<b>135</b>	<b>135</b>		<b>137</b>		± 5	<b>178</b>	<b>179</b>		<b>182</b>		± 7
		mmol/l	<b>3.79</b>	<b>3.73</b>		<b>3.60</b>		± 0.25	<b>8.38</b>	<b>8.38</b>		<b>8.51</b>		± 0.31	<b>11.05</b>	<b>11.12</b>		<b>11.30</b>		± 0.44
HT	HCT	%	<b>16.5</b>	<b>16.4</b>		<b>16.1</b>		± 2.0	<b>37.3</b>	<b>36.6</b>		<b>37.1</b>		± 2.5	<b>49.4</b>	<b>49.3</b>		<b>49.2</b>		± 3.0
		l/l	<b>0.165</b>	<b>0.164</b>		<b>0.161</b>		± 0.020	<b>0.373</b>	<b>0.366</b>		<b>0.371</b>		± 0.025	<b>0.494</b>	<b>0.493</b>		<b>0.492</b>		± 0.030
VGM	MCV	µm <sup>3</sup> ; fl	<b>70</b>	<b>70</b>		<b>69.5</b>		± 4	<b>81</b>	<b>80</b>		<b>80.5</b>		± 4	<b>87</b>	<b>86</b>		<b>85.5</b>		± 4
TGMH	MCH	pg	<b>26.0</b>	<b>25.6</b>		<b>25.1</b>		± 2.4	<b>29.3</b>	<b>29.5</b>		<b>29.7</b>		± 2.4	<b>31.3</b>	<b>31.2</b>		<b>31.6</b>		± 2.8
		fmol	<b>1.61</b>	<b>1.59</b>		<b>1.56</b>		± 0.15	<b>1.82</b>	<b>1.83</b>		<b>1.85</b>		± 0.15	<b>1.95</b>	<b>1.94</b>		<b>1.96</b>		± 0.18
CCMH	MCHC	g/dl	<b>37.1</b>	<b>36.6</b>		<b>36.1</b>		± 3.0	<b>36.2</b>	<b>36.8</b>		<b>36.9</b>		± 3.0	<b>36.0</b>	<b>36.3</b>		<b>37.0</b>		± 3.0
		g/l	<b>371</b>	<b>366</b>		<b>361</b>		± 30	<b>362</b>	<b>368</b>		<b>369</b>		± 30	<b>360</b>	<b>363</b>		<b>370</b>		± 30
		mmol/l	<b>23.03</b>	<b>22.75</b>		<b>22.43</b>		± 1.86	<b>22.45</b>	<b>22.88</b>		<b>22.93</b>		± 1.86	<b>22.37</b>	<b>22.56</b>		<b>22.95</b>		± 1.86
IDR	RDW	%	<b>14.5</b>	<b>14.5</b>		<b>16.0</b>		± 3.0	<b>13.0</b>	<b>13.0</b>		<b>13.0</b>		± 3.0	<b>12.0</b>	<b>12.5</b>		<b>12.0</b>		± 3.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>76</b>	<b>80</b>		<b>69</b>		± 20	<b>258</b>	<b>256</b>		<b>254</b>		± 40	<b>495</b>	<b>510</b>		<b>511</b>		± 60
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.0</b>	<b>8.7</b>		<b>9.5</b>		± 2.0	<b>8.2</b>	<b>8.4</b>		<b>9.0</b>		± 2.0	<b>8.0</b>	<b>8.2</b>		<b>8.4</b>		± 2.0
LYMPHO		#	<b>1.4</b>	<b>1.4</b>		N/A		± 0.2	<b>2.6</b>	<b>2.6</b>		N/A		± 0.5	<b>3.1</b>	<b>3.1</b>		N/A		± 1.1
		%	<b>64.8</b>	<b>63.3</b>		N/A		± 8.0	<b>34.8</b>	<b>33.5</b>		N/A		± 6.0	<b>16.3</b>	<b>15.2</b>		N/A		± 5.0
MONO		#	<b>0.2</b>	<b>0.2</b>		N/A		± 0.2	<b>0.7</b>	<b>0.7</b>		N/A		± 0.4	<b>1.8</b>	<b>1.7</b>		N/A		± 1.0
		%	<b>8.6</b>	<b>9.3</b>		N/A		± 7.0	<b>8.8</b>	<b>9.2</b>		N/A		± 5.0	<b>8.8</b>	<b>8.7</b>		N/A		± 5.0
GRANULO		#	<b>0.5</b>	<b>0.5</b>		N/A		± 0.2	<b>4.3</b>	<b>4.3</b>		N/A		± 0.5	<b>15.2</b>	<b>15.4</b>		N/A		± 1.5
		%	<b>26.6</b>	<b>27.5</b>		N/A		± 7.0	<b>56.4</b>	<b>57.4</b>		N/A		± 6.0	<b>74.9</b>	<b>76.1</b>		N/A		± 7.0
CRP		mg/dl	<b>0.46</b>	<b>0.49</b>		<b>0.48</b>		± 0.40	<b>2.30</b>	<b>2.41</b>		<b>2.22</b>		± 0.80	<b>4.24</b>	<b>4.35</b>		<b>4.18</b>		± 1.50
		mg/l	<b>4.6</b>	<b>4.9</b>		<b>4.8</b>		± 4.0	<b>23.0</b>	<b>24.1</b>		<b>22.2</b>		± 8.0	<b>42.4</b>	<b>43.5</b>		<b>41.8</b>		± 15.0

Ref: TEMP-1070 Rev.42 BACK/VERSO 1300092170