



Yumizen G DDi 2 for YUMIZEN G1500/G1550



Reference No.:		1300036391	
Lot No.:		930706 (rev. 1)	
Expiry date:		2024-07-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G1500/G1550	µg(FEU)/mL	ΔOD
		4,00	0,2580
		3,00	0,1945
		2,00	0,1220
		1,00	0,0555
		0,50	0,0264
		0,25	0,0130
Slope value		0,06	

Yumizen G DDi 2 for YUMIZEN G800



Reference No.:		1300036391	
Lot No.:		930706 (rev. 1)	
Expiry date:		2024-07-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G800	µg(FEU)/mL	ΔOD
		4,00	0,2530
		3,00	0,1952
		2,00	0,1219
		1,00	0,0525
		0,50	0,0242
		0,25	0,0118
Slope value		0,06	



Yumizen G DDi 2 for YUMIZEN G400 DDi/G200

Reference No.:		1300036391	
Lot No.:		930706 (rev. 1)	
Expiry date:		2024-07-31	
Coagulometers		MASTER CURVE	
		µg(FEU)/mL	ΔOD
Optical	Yumizen G400 DDi/G200	4,00	0,0990
		3,00	0,0730
		2,00	0,0475
		1,00	0,0231
		0,50	0,0115
		0,25	0,0059
		Slope value	0,02
Important notice			
<p>For Yumizen G400 DDi and Yumizen G200, make sure that the end time is set to 180 seconds in the D-Dimer setup and the curve type Lin-Lin p-p is selected in Calibration > D-Dimer > ug FEU/ml. For more information, please contact your local HORIBA Medical representative.</p>			