



# Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 2 0 9 3 3 4 6 4 1 2 0 9 2 4 D



D 5 1 0 9 1 4 8 0 7 1 1 1 8 0 8 2 D



D 6 1 2 3 1 0 1 1 3 7 0 7 9 1 3 1 D



D 7 0 7 8 1 1 7 0 8 8 1 1 9 0 7 3 D



D 8 1 2 2 0 8 1 1 2 1 3 4 7 4 7 0 D



D 9 2 7 6 3 7 3 2 0 0 3 0 0 2 2 6 D



D 0 3 3 9 3 5 0 D

Reference No.:		1300036412	
Lot No:		920933 (rev. 3)	
Expiry date:		2024-09-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,9-14,8
	Prothrombin %	%	71-118
	Prothrombin INR	INR	0,82-1,23
PT Liq (*)	Prothrombin Time	sec	10,1-13,7 (*)
	Prothrombin %	%	79-131 (*)
	Prothrombin INR	INR	0,78-1,17 (*)
PT R	Prothrombin Time	sec	8,8-11,9
	Prothrombin %	%	73-122
	Prothrombin INR	INR	0,81-1,21
APTT	Partial Thromboplastin Time	sec	34,7-47,0
APTT Liq	Partial Thromboplastin Time	sec	27,6-37,3
FIB Cl	Fibrinogen	g/L	2,0-3,0
TT	Thrombin Time	sec	22,6-33,9
AT	Antithrombin	%	92-125

### Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.

(\*) Warning! This range has to be used only with Yumizen G PT Liq 4 lots : 920315 and higher except for the Yumizen G PT Liq 4 lots 920611 & 920705. For the other lots, including 920611 & 920705, please refer to the previous revision of values sheet.



# Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 2 0 9 3 3 5 6 4 1 2 0 9 2 4 D



D 5 1 3 9 2 0 8 0 3 4 0 5 7 1 1 6 D



D 6 1 7 4 1 3 4 2 0 1 0 3 8 0 6 3 D



D 7 1 2 0 1 8 0 1 1 6 1 7 5 0 4 0 D



D 8 0 6 7 1 1 3 1 6 9 5 6 7 7 6 7 D



D 9 5 7 8 7 8 2 0 7 0 1 5 0 0 0 0 D



D 0 9 9 9 3 7 7 D

Reference No.:		1300036412	
Lot No.:		920933 (rev. 3)	
Expiry date:		2024-09-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	13,9-20,8
	Prothrombin %	%	34-57
	Prothrombin INR	INR	1,16-1,74
PT Liq (*)	Prothrombin Time	sec	13,4-20,1 (*)
	Prothrombin %	%	38-63 (*)
	Prothrombin INR	INR	1,20-1,80 (*)
PT R	Prothrombin Time	sec	11,6-17,5
	Prothrombin %	%	40-67
	Prothrombin INR	INR	1,13-1,69
APTT	Partial Thromboplastin Time	sec	56,7-76,7
APTT Liq	Partial Thromboplastin Time	sec	57,8-78,2
FIB Cl	Fibrinogen	g/L	0,7-1,5
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	29-48

### Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.

(\*) Warning! This range has to be used only with Yumizen G PT Liq 4 lots : 920315 and higher except for the Yumizen G PT Liq 4 lots 920611 & 920705. For the other lots, including 920611 & 920705, please refer to the previous revision of values sheet.