

ABX Minocal

LOT

CX 488

CAL



(Exp.) 2023-12-05
(YYYY-MM-DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|--------------|--------------|--------|--------|-----------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P 60 18P | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.2 | 8.9 | | | 9.3 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.42 | 4.35 | | | 4.42 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.1 | 12.8 | | | 13.1 | | | | | ± 0.2 |
| | g/l | 131 | 128 | | | 131 | | | | | ± 2 |
| | mmol/l | 8.14 | 7.97 | | | 8.14 | | | | | ± 0.12 |
| HT HCT | % | 36.7 | 36.8 | | | 37.3 | | | | | ± 1.0 |
| | l/l | 0.367 | 0.368 | | | 0.373 | | | | | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | 227 | 240 | | | 236 | | | | | ± 10 |
| VMP MPV | µm ³ ; fl | 8.9 | 8.7 | | | 9.0 | | | | | ± 0.5 |

ABX Minocal

LOT

CX 488

Rev 1

CAL



(Exp.) **2023-12-05**
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | LyseVet | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|-----------------------------|--------|--|--|--|--|--|-------------------------|
| | | MICROS | | MICROS | | | | | | |
| | | Scil Vet ABC | Scil Vet ABC Plus ESV 60 | | | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.0 | 8.9 | | | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.40 | 4.39 | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.0 | 12.9 | | | | | | | ± 0.2 |
| | g/l | 130 | 129 | | | | | | | ± 2 |
| | mmol/l | 8.07 | 8.01 | | | | | | | ± 0.12 |
| HT HCT | % | 37.4 | 37.1 | | | | | | | ± 1.0 |
| | l/l | 0.374 | 0.371 | | | | | | | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | 219 | 231 | | | | | | | ± 10 |
| VMP MPV | µm ³ ; fl | N/A | 9.0 | | | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.50 FRONT / RECTO 1300113555