



**LOT**

**MX 444 L**

**CONTROL L**



**2024-02-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.9	1.9			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.23	2.23			2.24					± 0.17
HB HGB	g/dl	5.7	5.7			5.5					± 0.5
	g/l	57	57			55					± 5
	mmol/l	3.54	3.53			3.41					± 0.31
HT HCT	%	15.8	15.6			15.9					± 2.0
	l/l	0.158	0.156			0.159					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	71	70			71					± 4
TGMH MCH	pg	25.6	25.5			24.5					± 2.0
	fmol	1.59	1.58			1.52					± 0.12
CCMH MCHC	g/dl	36.0	36.4			34.5					± 3.0
	g/l	360	364			345					± 30
	mmol/l	22.36	22.61			21.42					± 1.86
IDR RDW	%	15.0	13.5			15.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	74			77					± 22
VPM MPV	µm <sup>3</sup> ; fl	8.7	8.5			7.9					± 2.0
LYMPHO	#	1.20	1.23			1.29					± 0.40
	%	66.0	66.2			67.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	7.5	9.0			7.5					± 6.0
GRANULO	#	0.50	0.47			0.51					± 0.40
	%	26.5	25.0			25.5					± 7.0

<b>LOT</b>		<b>MX 444 N</b>	<b>CONTROL N</b>				(Exp.)				<b>2024-02-05</b> (YYYY-MM-DD)
Rev 1		<b>Minilysebio</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>7.4</b>	<b>7.2</b>			<b>7.2</b>					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.55</b>	<b>4.51</b>			<b>4.55</b>					± 0.26
HB HGB	g/dl	<b>13.5</b>	<b>13.3</b>			<b>13.2</b>					± 0.7
	g/l	<b>135</b>	<b>133</b>			<b>132</b>					± 7
	mmol/l	<b>8.38</b>	<b>8.26</b>			<b>8.22</b>					± 0.43
HT HCT	%	<b>36.9</b>	<b>36.5</b>			<b>36.9</b>					± 2.7
	l/l	<b>0.369</b>	<b>0.365</b>			<b>0.369</b>					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	<b>81</b>	<b>81</b>			<b>81</b>					± 4
TGMH MCH	pg	<b>29.7</b>	<b>29.5</b>			<b>29.1</b>					± 2.0
	fmol	<b>1.84</b>	<b>1.83</b>			<b>1.81</b>					± 0.12
CCMH MCHC	g/dl	<b>36.6</b>	<b>36.4</b>			<b>35.9</b>					± 3.0
	g/l	<b>366</b>	<b>364</b>			<b>359</b>					± 30
	mmol/l	<b>22.75</b>	<b>22.61</b>			<b>22.29</b>					± 1.86
IDR RDW	%	<b>14.0</b>	<b>13.5</b>			<b>14.5</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>243</b>	<b>249</b>			<b>256</b>					± 40
VPM MPV	µm <sup>3</sup> ; fl	<b>8.1</b>	<b>7.9</b>			<b>7.3</b>					± 2.0
LYMPHO	#	<b>2.63</b>	<b>2.56</b>			<b>2.56</b>					± 0.40
	%	<b>35.5</b>	<b>35.5</b>			<b>35.5</b>					± 5.0
MONO	#	<b>0.48</b>	<b>0.51</b>			<b>0.43</b>					± 0.30
	%	<b>6.5</b>	<b>7.1</b>			<b>6.0</b>					± 4.0
GRANULO	#	<b>4.29</b>	<b>4.13</b>			<b>4.21</b>					± 0.60
	%	<b>58.0</b>	<b>57.3</b>			<b>58.5</b>					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

<b>LOT</b>		<b>MX 444 H</b>		<b>CONTROL H</b>								(Exp.) <b>2024-02-05</b> (YYYY-MM-DD)				
		Rev 1		<b>Milynlysebio</b>												
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200					60 8P - 60 18P								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>19.9</b>	<b>19.4</b>					<b>19.5</b>								± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.66</b>	<b>5.62</b>					<b>5.71</b>								± 0.32
HB HGB	g/dl	<b>18.3</b>	<b>17.8</b>					<b>17.9</b>								± 0.9
	g/l	<b>183</b>	<b>178</b>					<b>179</b>								± 9
	mmol/l	<b>11.36</b>	<b>11.05</b>					<b>11.13</b>								± 0.56
HT HCT	%	<b>50.4</b>	<b>49.5</b>					<b>50.2</b>								± 3.4
	l/l	<b>0.504</b>	<b>0.495</b>					<b>0.502</b>								± 0.034
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>88</b>					<b>88</b>								± 4
TGMH MCH	pg	<b>32.3</b>	<b>31.7</b>					<b>31.4</b>								± 2.0
	fmol	<b>2.01</b>	<b>1.97</b>					<b>1.95</b>								± 0.12
CCMH MCHC	g/dl	<b>36.3</b>	<b>36.0</b>					<b>35.7</b>								± 3.0
	g/l	<b>363</b>	<b>360</b>					<b>357</b>								± 30
	mmol/l	<b>22.56</b>	<b>22.35</b>					<b>22.15</b>								± 1.86
IDR RDW	%	<b>14.0</b>	<b>13.5</b>					<b>14.5</b>								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>503</b>	<b>516</b>					<b>518</b>								± 65
VPM MPV	µm <sup>3</sup> ; fl	<b>8.1</b>	<b>7.8</b>					<b>7.4</b>								± 2.0
LYMPHO	#	<b>3.38</b>	<b>3.29</b>					<b>3.22</b>								± 1.00
	%	<b>17.0</b>	<b>17.0</b>					<b>16.5</b>								± 5.0
MONO	#	<b>0.90</b>	<b>1.36</b>					<b>0.98</b>								± 0.80
	%	<b>4.5</b>	<b>7.0</b>					<b>5.0</b>								± 4.0
GRANULO	#	<b>15.62</b>	<b>14.73</b>					<b>15.31</b>								± 1.40
	%	<b>78.5</b>	<b>76.0</b>					<b>78.5</b>								± 7.0