


**LOT** PX 444

**CONTROL**

 (Exp.) **2024-01-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE				
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE	
		PENTRA						PENTRA						PENTRA						
			DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS						
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>3.0</b>	<b>3.0</b>			± 0.4	<b>8.9</b>	<b>8.9</b>			± 1.0	<b>19.4</b>	<b>19.4</b>			± 2.2				
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.32</b>	<b>2.32</b>			± 0.16	<b>4.68</b>	<b>4.68</b>			± 0.20	<b>5.20</b>	<b>5.20</b>			± 0.25				
HB HGB	g/dl	<b>6.3</b>	<b>6.3</b>			± 0.4	<b>13.6</b>	<b>13.6</b>			± 0.5	<b>16.2</b>	<b>16.2</b>			± 0.6				
	g/l	<b>63</b>	<b>63</b>			± 4	<b>136</b>	<b>136</b>			± 5	<b>162</b>	<b>162</b>			± 6				
	mmol/l	<b>3.91</b>	<b>3.91</b>			± 0.25	<b>8.45</b>	<b>8.45</b>			± 0.31	<b>10.06</b>	<b>10.06</b>			± 0.37				
HT HCT	%	<b>19.0</b>	<b>19.0</b>			± 1.5	<b>39.8</b>	<b>39.8</b>			± 2.0	<b>46.8</b>	<b>46.8</b>			± 2.5				
	l/l	<b>0.190</b>	<b>0.190</b>			± 0.015	<b>0.398</b>	<b>0.398</b>			± 0.020	<b>0.468</b>	<b>0.468</b>			± 0.025				
VGM MCV	µm <sup>3</sup> ; fl	<b>82</b>	<b>82</b>			± 5	<b>85</b>	<b>85</b>			± 5	<b>90</b>	<b>90</b>			± 5				
TGMH MCH	pg	<b>27.2</b>	<b>27.2</b>			± 2.0	<b>29.1</b>	<b>29.1</b>			± 2.0	<b>31.2</b>	<b>31.2</b>			± 2.5				
	fmol	<b>1.69</b>	<b>1.69</b>			± 0.12	<b>1.80</b>	<b>1.80</b>			± 0.12	<b>1.93</b>	<b>1.93</b>			± 0.16				
CCMH MCHC	g/dl	<b>33.1</b>	<b>33.1</b>			± 3.0	<b>34.2</b>	<b>34.2</b>			± 3.0	<b>34.6</b>	<b>34.6</b>			± 3.0				
	g/l	<b>331</b>	<b>331</b>			± 30	<b>342</b>	<b>342</b>			± 30	<b>346</b>	<b>346</b>			± 30				
	mmol/l	<b>20.57</b>	<b>20.57</b>			± 1.86	<b>21.23</b>	<b>21.23</b>			± 1.86	<b>21.50</b>	<b>21.50</b>			± 1.86				
IDR RDW	%	<b>17.0</b>	<b>17.0</b>			± 4.0	<b>14.0</b>	<b>14.0</b>			± 4.0	<b>13.0</b>	<b>13.0</b>			± 4.0				
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>76</b>	<b>76</b>			± 20	<b>238</b>	<b>238</b>			± 30	<b>480</b>	<b>480</b>			± 50				
VPM MPV	µm <sup>3</sup> ; fl	<b>8.9</b>	<b>8.9</b>			± 2.0	<b>8.6</b>	<b>8.6</b>			± 2.0	<b>9.6</b>	<b>9.6</b>			± 2.0				
NEUT	#	<b>1.80</b>	<b>1.80</b>			± 0.35	<b>5.37</b>	<b>5.37</b>			± 0.90	<b>14.10</b>	<b>14.10</b>			± 1.90				
	%	<b>60.0</b>	<b>60.0</b>			± 10.0	<b>60.3</b>	<b>60.3</b>			± 10.0	<b>72.7</b>	<b>72.7</b>			± 10.0				
LYMPHO	#	<b>0.76</b>	<b>0.76</b>			± 0.33	<b>2.53</b>	<b>2.53</b>			± 0.70	<b>3.01</b>	<b>3.01</b>			± 1.50				
	%	<b>25.4</b>	<b>25.4</b>			± 12.0	<b>28.4</b>	<b>28.4</b>			± 8.0	<b>15.5</b>	<b>15.5</b>			± 8.0				
MONO	#	<b>0.11</b>	<b>0.11</b>			± 0.11	<b>0.27</b>	<b>0.27</b>			± 0.27	<b>0.39</b>	<b>0.39</b>			± 0.39				
	%	<b>3.6</b>	<b>3.6</b>			± 3.6	<b>3.0</b>	<b>3.0</b>			± 3.0	<b>2.0</b>	<b>2.0</b>			± 2.0				
EOS	#	<b>0.20</b>	<b>0.20</b>			± 0.20	<b>0.37</b>	<b>0.37</b>			± 0.37	<b>1.01</b>	<b>1.01</b>			± 1.01				
	%	<b>6.6</b>	<b>6.6</b>			± 6.6	<b>4.2</b>	<b>4.2</b>			± 4.2	<b>5.2</b>	<b>5.2</b>			± 5.2				
BASO	#	<b>0.13</b>	<b>0.13</b>			± 0.13	<b>0.36</b>	<b>0.36</b>			± 0.36	<b>0.89</b>	<b>0.89</b>			± 0.89				
	%	<b>4.4</b>	<b>4.4</b>			± 4.4	<b>4.1</b>	<b>4.1</b>			± 4.1	<b>4.6</b>	<b>4.6</b>			± 4.6				

Ref: TEMP-0821 Rev.49 BACK / VERSO