

# ABX Minotrol Retic



**LOT** RX 444  
Rev 1

**CONTROL**

(Exp.) 2024-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.6	1.6	± 0.8	4.9	4.9	± 2.0	10.5	10.5	± 3.2			
RETIC	%	1.53	1.53	± 0.77	4.83	4.83	± 1.21	10.06	10.06	± 2.01			
RETIC	10 <sup>6</sup> /µl	0.052	0.052	± 0.026	0.165	0.165	± 0.041	0.341	0.341	± 0.068			
MFI MRI *	%	32.3	32.3	± 8.0	47.9	47.9	± 10.0	52.7	52.7	± 12.0			
GR RBC	10 <sup>6</sup> /µl	3.39	3.39	± 0.30	3.42	3.42	± 0.30	3.39	3.39	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.44 FRONT / RECTO 1300090305

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 444  
Rev 1

**CONTROL**

(Exp.) 2024-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	<b>1.6</b>			± 0.8	<b>4.9</b>			± 2.0	<b>10.5</b>			± 3.2
RETIC	%	<b>1.69</b>			± 0.85	<b>5.15</b>			± 1.29	<b>10.55</b>			± 2.11
RETIC	10 <sup>6</sup> /µl	<b>0.057</b>			± 0.029	<b>0.174</b>			± 0.044	<b>0.352</b>			± 0.070
MFI MRI *	%	<b>31.1</b>			± 8.0	<b>42.5</b>			± 10.0	<b>45.7</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.35</b>			± 0.30	<b>3.37</b>			± 0.30	<b>3.34</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.44 BACK / VERSO 1300090305

\* For research use only in Canada and USA