

LOT

MX 444 L  
Rev 1

CONTROL

L

 (Exp.)

2024-02-05  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l			2.0						± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l			2.28						± 0.15
HB HGB	g/dl			6.0						± 0.4
	g/l			60						± 4
	mmol/l			3.73						± 0.25
HT HCT	%			16.0						± 2.0
	l/l			0.160						± 0.020
VGM MCV	µm <sup>3</sup> ; fl			70						± 4
TGMH MCH	pg			26.3						± 2.0
	fmol			1.63						± 0.12
CCMH MCHC	g/dl			37.6						± 3.0
	g/l			376						± 30
	mmol/l			23.35						± 1.86
IDR RDW	%			14.0						± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l			69						± 20
VPM MPV	µm <sup>3</sup> ; fl			8.0						± 2.0
LYMPHO	#			1.24						± 0.40
	%			64.5						± 8.0
MONO	#			0.20						± 0.20
	%			7.5						± 6.0
GRANULO	#			0.56						± 0.40
	%			28.0						± 7.0

**LOT**

**MX 444 N**  
Rev 1

**CONTROL**

**N**

(Exp.)

**2024-02-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l			7.6						± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l			4.52						± 0.18
HB HGB	g/dl			13.4						± 0.5
	g/l			134						± 5
	mmol/l			8.32						± 0.31
HT HCT	%			36.6						± 2.5
	l/l			0.366						± 0.025
VGM MCV	µm <sup>3</sup> ; fl			81						± 4
TGMH MCH	pg			29.6						± 2.0
	fmol			1.84						± 0.12
CCMH MCHC	g/dl			36.6						± 3.0
	g/l			366						± 30
	mmol/l			22.73						± 1.86
IDR RDW	%			13.5						± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l			251						± 40
VPM MPV	µm <sup>3</sup> ; fl			8.0						± 2.0
LYMPHO	#			2.62						± 0.40
	%			34.5						± 5.0
MONO	#			0.57						± 0.30
	%			7.5						± 4.0
GRANULO	#			4.41						± 0.60
	%			58.0						± 6.0

Ret: TEMP-0830 Rev.42 BACK / VERSO 1300104144

LOT

MX 444 H  
Rev 1

CONTROL

H

 (Exp.)

2024-02-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l			20.0						± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l			5.52						± 0.20
HB HGB	g/dl			17.7						± 0.6
	g/l			177						± 6
	mmol/l			10.99						± 0.37
HT HCT	%			48.6						± 3.0
	l/l			0.486						± 0.030
VGM MCV	µm <sup>3</sup> ; fl			88						± 4
TGMH MCH	pg			32.1						± 2.0
	fmol			1.99						± 0.12
CCMH MCHC	g/dl			36.4						± 3.0
	g/l			364						± 30
	mmol/l			22.63						± 1.86
IDR RDW	%			12.5						± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l			494						± 45
VPM MPV	µm <sup>3</sup> ; fl			8.2						± 2.0
LYMPHO	#			3.40						± 1.00
	%			17.0						± 5.0
MONO	#			1.20						± 0.80
	%			6.0						± 4.0
GRANULO	#			15.40						± 1.40
	%			77.0						± 7.0