


**LOT** PX 444

**CONTROL**

 (Exp.) 2024-01-05  
(YYYY-MM-DD)

Rev 1

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120	DX NEXUS DF NEXUS						DX120 DF120		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.0	3.0			± 0.4	8.9	8.9			± 1.0	19.4	19.4			± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.32	2.32			± 0.16	4.68	4.68			± 0.20	5.20	5.20			± 0.25			
HB HGB	g/dl	6.3	6.3			± 0.4	13.6	13.6			± 0.5	16.2	16.2			± 0.6			
	g/l	63	63			± 4	136	136			± 5	162	162			± 6			
	mmol/l	3.91	3.91			± 0.25	8.45	8.45			± 0.31	10.06	10.06			± 0.37			
HT HCT	%	19.0	19.0			± 1.5	39.8	39.8			± 2.0	46.8	46.8			± 2.5			
	l/l	0.190	0.190			± 0.015	0.398	0.398			± 0.020	0.468	0.468			± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	82	82			± 5	85	85			± 5	90	90			± 5			
TGMH MCH	pg	27.2	27.2			± 2.0	29.1	29.1			± 2.0	31.2	31.2			± 2.5			
	fmol	1.69	1.69			± 0.12	1.80	1.80			± 0.12	1.93	1.93			± 0.16			
CCMH MCHC	g/dl	33.1	33.1			± 3.0	34.2	34.2			± 3.0	34.6	34.6			± 3.0			
	g/l	331	331			± 30	342	342			± 30	346	346			± 30			
	mmol/l	20.57	20.57			± 1.86	21.23	21.23			± 1.86	21.50	21.50			± 1.86			
IDR RDW	%	17.0	17.0			± 4.0	14.0	14.0			± 4.0	13.0	13.0			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	76	76			± 20	238	238			± 30	480	480			± 50			
VPM MPV	µm <sup>3</sup> ; fl	8.9	8.9			± 2.0	8.6	8.6			± 2.0	9.6	9.6			± 2.0			
NEUT	#	1.80	1.80			± 0.35	5.37	5.37			± 0.90	14.10	14.10			± 1.90			
	%	60.0	60.0			± 10.0	60.3	60.3			± 10.0	72.7	72.7			± 10.0			
LYMPHO	#	0.76	0.76			± 0.33	2.53	2.53			± 0.70	3.01	3.01			± 1.50			
	%	25.4	25.4			± 12.0	28.4	28.4			± 8.0	15.5	15.5			± 8.0			
MONO	#	0.11	0.11			± 0.11	0.27	0.27			± 0.27	0.39	0.39			± 0.39			
	%	3.6	3.6			± 3.6	3.0	3.0			± 3.0	2.0	2.0			± 2.0			
EOS	#	0.20	0.20			± 0.20	0.37	0.37			± 0.37	1.01	1.01			± 1.01			
	%	6.6	6.6			± 6.6	4.2	4.2			± 4.2	5.2	5.2			± 5.2			
BASO	#	0.13	0.13			± 0.13	0.36	0.36			± 0.36	0.89	0.89			± 0.89			
	%	4.4	4.4			± 4.4	4.1	4.1			± 4.1	4.6	4.6			± 4.6			

Ref: TEMP-0821 Rev.49 BACK / VERSO