


LOT PX 444
Rev 1


CONTROL

 (Exp.) 2024-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															
			CONTROL				L	CONTROL				N	CONTROL				H	TOLERANCES TOLERANCE
			PENTRA					PENTRA					PENTRA					
			XL80	XLR				XL80	XLR				XL80	XLR				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0			± 0.4	8.7	8.7			± 1.0	18.8	18.8			± 2.2	
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.28	2.28			± 0.16	4.59	4.59			± 0.20	5.08	5.08			± 0.25	
HB	HGB	g/dl	6.2	6.2			± 0.4	13.6	13.6			± 0.5	16.2	16.2			± 0.6	
		g/l	62	62			± 4	136	136			± 5	162	162			± 6	
		mmol/l	3.85	3.85			± 0.25	8.45	8.45			± 0.31	10.06	10.06			± 0.37	
HT	HCT	%	18.5	18.5			± 1.5	39.0	39.0			± 2.0	45.7	45.7			± 2.5	
		l/l	0.185	0.185			± 0.015	0.390	0.390			± 0.020	0.457	0.457			± 0.025	
VGM	MCV	µm ³ ; fl	81	81			± 5	85	85			± 5	90	90			± 5	
TGMH	MCH	pg	27.2	27.2			± 2.0	29.6	29.6			± 2.0	31.9	31.9			± 2.5	
		fmol	1.69	1.69			± 0.12	1.84	1.84			± 0.12	1.98	1.98			± 0.16	
CCMH	MCHC	g/dl	33.6	33.6			± 3.0	34.9	34.9			± 3.0	35.4	35.4			± 3.0	
		g/l	336	336			± 30	349	349			± 30	354	354			± 30	
		mmol/l	20.85	20.85			± 1.86	21.65	21.65			± 1.86	22.00	22.00			± 1.86	
IDR	RDW	%	14.5	14.5			± 4.0	13.0	13.0			± 4.0	12.0	12.0			± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	71	71			± 20	232	232			± 30	462	462			± 50	
VPM	MPV	µm ³ ; fl	8.7	8.7			± 2.0	9.0	9.0			± 2.0	9.9	9.9			± 2.0	
NEUT		#	1.65	1.65			± 0.35	4.81	4.81			± 0.90	13.35	13.35			± 1.90	
		%	54.9	54.9			± 10.0	55.3	55.3			± 10.0	71.0	71.0			± 10.0	
LYMPHO		#	0.99	0.99			± 0.33	3.08	3.08			± 0.70	3.52	3.52			± 1.50	
		%	32.9	32.9			± 12.0	35.4	35.4			± 8.0	18.7	18.7			± 8.0	
MONO		#	0.09	0.09			± 0.09	0.25	0.25			± 0.25	0.39	0.39			± 0.39	
		%	2.9	2.9			± 2.9	2.9	2.9			± 2.9	2.1	2.1			± 2.1	
EOS		#	0.19	0.19			± 0.19	0.30	0.30			± 0.30	0.81	0.81			± 0.81	
		%	6.4	6.4			± 6.4	3.5	3.5			± 3.5	4.3	4.3			± 4.3	
BASO		#	0.09	0.09			± 0.09	0.25	0.25			± 0.25	0.73	0.73			± 0.73	
		%	2.9	2.9			± 2.9	2.9	2.9			± 2.9	3.9	3.9			± 3.9	

LOT PX 444
Rev 1

CONTROL

 (Exp.) 2024-01-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0	3.0		± 0.4	8.7	8.7	8.5		± 1.0	19.2	19.0	18.6		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.32	2.32	2.30		± 0.16	4.62	4.59	4.57		± 0.20	5.09	5.09	5.05		± 0.25			
	g/dl	6.1	6.2	6.2		± 0.4	13.6	13.5	13.5		± 0.5	16.2	16.0	16.1		± 0.6			
HB HGB	g/l	61	62	62		± 4	136	135	135		± 5	162	160	161		± 6			
	mmol/l	3.79	3.85	3.85		± 0.25	8.45	8.38	8.38		± 0.31	10.06	9.94	10.00		± 0.37			
HT HCT	%	18.1	18.1	18.1		± 1.5	39.3	38.6	38.6		± 2.0	46.8	45.8	45.7		± 2.5			
	l/l	0.181	0.181	0.181		± 0.015	0.393	0.386	0.386		± 0.020	0.468	0.458	0.457		± 0.025			
VGM MCV	µm ³ ·fl	78	78	78.5		± 5	85	84	84.5		± 5	92	90	90.5		± 5			
TGMH MCH	pg	26.3	26.7	27.0		± 2.0	29.4	29.4	29.5		± 2.0	31.8	31.4	31.9		± 2.5			
	fmol	1.63	1.66	1.67		± 0.12	1.83	1.83	1.83		± 0.12	1.98	1.95	1.98		± 0.16			
	g/dl	33.7	34.3	34.3		± 3.0	34.6	35.0	35.0		± 3.0	34.6	34.9	35.2		± 3.0			
CCMH MCHC	g/l	337	343	343		± 30	346	350	350		± 30	346	349	352		± 30			
	mmol/l	20.93	21.28	21.32		± 1.86	21.51	21.74	21.71		± 1.86	21.48	21.69	21.88		± 1.86			
IDR RDW	%	14.5	14.5	13.5		± 4.0	13.0	13.0	11.5		± 4.0	12.0	12.0	11.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	73	71	70		± 20	229	228	219		± 30	465	468	442		± 50			
VPM MPV	µm ³ ·fl	8.5	8.6	8.1		± 2.0	8.3	8.4	8.0		± 2.0	9.0	9.1	8.7		± 2.0			
NEUT	#	1.65	1.61	1.61		± 0.35	4.84	4.79	4.63		± 0.90	13.52	13.51	13.21		± 1.90			
	%	55.0	53.8	53.5		± 10.0	55.6	55.1	54.5		± 10.0	70.4	71.1	71.0		± 10.0			
LYMPHO	#	0.99	1.03	1.07		± 0.33	3.02	3.10	3.10		± 0.70	3.55	3.53	3.53		± 1.50			
	%	33.0	34.4	35.5		± 12.0	34.7	35.6	36.5		± 8.0	18.5	18.6	19.0		± 8.0			
MONO	#	0.09	0.08	0.09		± 0.08	0.25	0.24	0.26		± 0.24	0.40	0.32	0.28		± 0.28			
	%	2.9	2.5	3.0		± 2.5	2.9	2.8	3.0		± 2.8	2.1	1.7	1.5		± 1.5			
EOS	#	0.19	0.19	0.15		± 0.15	0.33	0.30	0.26		± 0.26	0.96	0.87	0.84		± 0.84			
	%	6.2	6.4	5.0		± 5.0	3.8	3.5	3.0		± 3.0	5.0	4.6	4.5		± 4.5			
BASO	#	0.09	0.09	0.09		± 0.09	0.26	0.26	0.26		± 0.26	0.77	0.76	0.74		± 0.74			
	%	2.9	2.9	3.0		± 2.9	3.0	3.0	3.0		± 3.0	4.0	4.0	4.0		± 4.0			

Ref: TEMP-0821 Rev.49 FRONT / RECTO