

ABX Minotrol Retic



LOT RX 444
Rev 1

CONTROL

(Exp.) 2024-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.6	1.6	± 0.8	4.9	4.9	± 2.0	10.5	10.5	± 3.2			
RETIC	%	1.53	1.53	± 0.77	4.83	4.83	± 1.21	10.06	10.06	± 2.01			
RETIC	10 ⁶ /µl	0.052	0.052	± 0.026	0.165	0.165	± 0.041	0.341	0.341	± 0.068			
MFI MRI *	%	32.3	32.3	± 8.0	47.9	47.9	± 10.0	52.7	52.7	± 12.0			
GR RBC	10 ⁶ /µl	3.39	3.39	± 0.30	3.42	3.42	± 0.30	3.39	3.39	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.44 FRONT / RECTO 1300090305

ABX Minotrol Retic



LOT RX 444
Rev 1

CONTROL

(Exp.) 2024-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.6			± 0.8	4.9			± 2.0	10.5			± 3.2
RETIC	%	1.69			± 0.85	5.15			± 1.29	10.55			± 2.11
RETIC	10 ⁶ /µl	0.057			± 0.029	0.174			± 0.044	0.352			± 0.070
MFI MRI *	%	31.1			± 8.0	42.5			± 10.0	45.7			± 12.0
GR RBC	10 ⁶ /µl	3.35			± 0.30	3.37			± 0.30	3.34			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.44 BACK / VERSO 1300090305

* For research use only in Canada and USA