



HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée
 B.P. 7290
 34184 Montpellier Cedex 4
 FRANCE
 www.horiba.com




LOT 23040 (rev. 01)



2 3 0 4 0

SAP 1300029568



1300029568

REF : 1300023947



2025/02/28

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400 (P400)	Pentra C400 (PC400)	Pentra C200 (PC200)	Unité	Valeur titrée	Intervalle de confiance	
				Unit	Assigned value	Confidence range	1 SD
C				Unité	Assigned value	Confidence range*	1 SD
AlbT2-U ABX Pentra Micro ALBUMIN 2 CP (1300032563)	•			mg/L mg/dL	61.7 6.17	49.3 - 74.1 4.93 - 7.41	6.2 0.62
		•		mg/L mg/dL	61.7 6.17	49.3 - 74.1 4.93 - 7.41	6.2 0.62
			•	mg/L mg/dL	57.6 5.76	46.0 - 69.2 4.60 - 6.92	5.8 0.58
AMY (EPS-G7) ABX Pentra Amylase CP (A11A01628)	•			U/L µkat/L	247.6 4.135	198.0 - 297.2 3.307 - 4.963	24.8 0.414
		•		U/L µkat/L	247.6 4.135	198.0 - 297.2 3.307 - 4.963	24.8 0.414
			•	U/L µkat/L	230.3 3.846	184.3 - 276.3 3.078 - 4.614	23.0 0.384
CA (AZ III) ABX Pentra Calcium AS CP (A11A01954)	•			mmol/L mg/L mg/dL	3.00 120.3 12.03	2.40 - 3.60 96.3 - 144.3 9.63 - 14.43	0.30 12.0 1.20
		•		mmol/L mg/L mg/dL	3.00 120.3 12.03	2.40 - 3.60 96.3 - 144.3 9.63 - 14.43	0.30 12.0 1.20
			•	mmol/L mg/L mg/dL	2.81 112.7 11.27	2.25 - 3.37 90.3 - 135.1 9.03 - 13.51	0.28 11.2 1.12
				mmol/L	3.00	2.40 - 3.60	0.30
				mg/L	120.3	96.3 - 144.3	12.0
				mg/dL	12.03	9.63 - 14.43	1.20

* Confidence range: +/- 2 SD

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD		
CREA (CREA3 / CREA4) (Jaffe/Alkaline Picrate) ABX Pentra Creatinine 120 CP (A11A01933)	•			μmol/L mmol/L mg/dL	13432 13.43 151.8	10746 - 16118 10.75 - 16.11 121.4 - 182.2	1343 1.34 15.2		
		•		μmol/L mmol/L mg/dL	13432 13.43 151.8	10746 - 16118 10.75 - 16.11 121.4 - 182.2	1343 1.34 15.2		
			•	μmol/L mmol/L mg/dL	12713 12.71 143.7	10171 - 15255 10.17 - 15.25 114.9 - 172.5	1271 1.27 14.4		
	CREA ENZ. ABX Pentra Enzymatic Creatinine CP (A11A01907)	•			μmol/L mmol/L mg/dL	14740 14.74 166.6	11792 - 17688 11.80 - 17.68 133.2 - 200.0	1474 1.47 16.7	
			•		μmol/L mmol/L mg/dL	14740 14.74 166.6	11792 - 17688 11.80 - 17.68 133.2 - 200.0	1474 1.47 16.7	
				•	μmol/L mmol/L mg/dL	14411 14.41 162.8	11529 - 17293 11.53 - 17.29 130.2 - 195.4	1441 1.44 16.3	
		GluK (HK) ABX Pentra Glucose HK CP (A11A01667)	•			mmol/L g/L mg/dL	16.35 2.943 294.3	13.07 - 19.63 2.353 - 3.533 235.3 - 353.3	1.64 0.295 29.5
				•		mmol/L g/L mg/dL	16.35 2.943 294.3	13.07 - 19.63 2.353 - 3.533 235.3 - 353.3	1.64 0.295 29.5
					•	mmol/L g/L mg/dL	16.40 2.952 295.2	13.12 - 19.68 2.362 - 3.542 236.2 - 354.2	1.64 0.295 29.5
GluP (PAP) ABX Pentra Glucose PAP CP (A11A01668)	•			mmol/L g/L mg/dL	18.57 3.343 334.3	14.85 - 22.29 2.673 - 4.013 267.3 - 401.3	1.86 0.335 33.5		
		•		mmol/L g/L mg/dL	18.57 3.343 334.3	14.85 - 22.29 2.673 - 4.013 267.3 - 401.3	1.86 0.335 33.5		
			•	mmol/L g/L mg/dL	16.53 2.975 297.5	13.23 - 19.83 2.381 - 3.569 238.1 - 356.9	1.65 0.297 29.7		
	PHOS (Ammonium Molybdate UV) ABX Pentra Phosphorus CP (A11A01665)	•			mmol/L mg/L mg/dL	16.64 516 51.6	13.32 - 19.96 414 - 618 41.4 - 61.8	1.66 51 5.1	
			•		mmol/L mg/L mg/dL	16.64 516 51.6	13.32 - 19.96 414 - 618 41.4 - 61.8	1.66 51 5.1	
				•	mmol/L mg/L mg/dL	16.04 497 49.7	12.84 - 19.24 397 - 597 39.7 - 59.7	1.60 50 5.0	

* Confidence range: +/- 2 SD

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD	
TPU (Pyrogallol red) ABX Pentra Urinary Proteins CP (A11A01642)	•			g/L	0.572	0.458 - 0.686	0.057	
				mg/L	572	458 - 686	57	
				mg/dL	57.2	45.8 - 68.6	5.7	
		•			g/L	0.572	0.458 - 0.686	0.057
					mg/L	572	458 - 686	57
					mg/dL	57.2	45.8 - 68.6	5.7
		•		g/L	0.599	0.479 - 0.719	0.060	
				mg/L	599	479 - 719	60	
				mg/dL	59.9	47.9 - 71.9	6.0	
UREA (GLDH) ABX Pentra Urea CP (A11A01641)	•			mmol/L	276	220 - 332	28	
				mg/L	16560	13200 - 19920	1680	
				mg/dL	1656	1320 - 1992	168	
		•			mmol/L	276	220 - 332	28
					mg/L	16560	13200 - 19920	1680
					mg/dL	1656	1320 - 1992	168
		•		mmol/L	246	196 - 296	25	
				mg/L	14760	11760 - 17760	1500	
				mg/dL	1476	1176 - 1776	150	
BUN (GLDH) ABX Pentra Urea CP (A11A01641)	•			mmol/L	276	220 - 332	28	
				mg/dL	775	617 - 933	79	
		•			mmol/L	276	220 - 332	28
					mg/dL	775	617 - 933	79
		•		mmol/L	246	196 - 296	25	
				mg/dL	691	551 - 831	70	
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	•			µmol/L	852	682 - 1022	85	
				mmol/L	0.85	0.67 - 1.03	0.09	
				mg/dL	14.31	11.45 - 17.17	1.43	
		•			µmol/L	852	682 - 1022	85
					mmol/L	0.85	0.67 - 1.03	0.09
					mg/dL	14.31	11.45 - 17.17	1.43
		•		µmol/L	840	672 - 1008	84	
				mmol/L	0.84	0.68 - 1.00	0.08	
				mg/dL	14.11	11.29 - 16.93	1.41	

* Confidence range: +/- 2 SD