

## Intended Use

Intended for use as an Lactate standard for standardization or calibration of Lactate Oxidase, Lactate procedures. **Rx Only**

## Summary and Explanation

The Lactate standard is prepared from lactic acid in an aqueous solution. The standard provides a concentration level of lactate within the linearity range of the Lactate Oxidase, Lactate procedure. Stabilizers and preservatives added.

## Principle

The standard should be used to calibrate the determination of unknown specimens using a Lactate Oxidase, Lactate assay. Controls should be included to validate the calibration of the assay.

## Ingredients

Reactive ingredients: A lactate solution prepared from lactic acid in an aqueous solution. Preservatives added. Consult vial label for assigned concentration for the standard.

Precautions: For *in vitro* diagnostic use. Handle with the same precautions used for human samples.

## Storage and stability

Store at 2-8°C. Stable until expiration date if sealed tightly and stored at 2-8°C.

## Preparation

Standard is provided as a ready to use liquid.

## Procedure

The standard should be used to standardize instrument or assay following the directions that accompany the instrument and reagent kit used.

Materials provided: An aqueous Lactate standard 5.0 mmol/L.










Materials not provided: Pointe Lactate Oxidase, Lactate reagent and controls.


Limitations: Factors that might cause inaccurate results are improper pipetting and inadequate mixing.

## Standard Value

The value is 5.0 mmol/L when used with the Pointe Lactate Oxidase, Lactate procedure.

## Symbol Key


 Use by (YYYY-MM-DD)	 Lot and batch code
 Catalog number	 Manufacturer
 <i>In vitro</i> diagnostic medical device	 Temperature limitation
 Consult instructions for use	<b>Rx Only:</b> Prescription Use Only
 CE mark	 Authorized representative in the European Community

 L7596-STD



Manufactured by  
HORIBA Instruments Incorporated  
5449 Research Drive  
Canton, MI 48188



Manufactured by HORIBA Instruments Incorporated - Pointe Brand 5449 Research Drive, Canton, MI 48188	
European Authorized Representative: Obelis s.a. Boulevard Général Wahis 53 1030 Brussels, BELGIUM Tel: (32)2.732.59.54 Fax:(32)2.732.60.03 email: mail@obelis.net	

## Certified to Perform Reagents

The Pointe reagents are certified to be manufactured according to specified parameters. Any Pointe reagent product not meeting specifications through its listed expiration date will be remedied immediately without charge.

Rev. 06/22 P803-L7596-05