

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT

MX 445 L
 Rev 1

CONTROL L

(Exp.)

2024-04-05
 (YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|----------------|--------|--------|--------|--------------------------|----------------|--------|--------|--------|-------------------------|
| | | MICROS ES60 | MICROS | MICROS | MICROS | MICROS 60 8P - 60 18P | ARGOS - HELIOS | PENTRA | PENTRA | PENTRA | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.0 | | | | 2.1 | | | | | ± 0.4 |
| GR RBC | 10 ⁹ /mm ³ ; 10 ¹² /l | 2.20 | | | | 2.16 | | | | | ± 0.17 |
| HB HGB | g/dl | 5.7 | | | | 5.5 | | | | | ± 0.5 |
| | g/l | 57 | | | | 55 | | | | | ± 5 |
| | mmol/l | 3.54 | | | | 3.41 | | | | | ± 0.31 |
| HT HCT | % | 15.6 | | | | 15.6 | | | | | ± 2.0 |
| | l/l | 0.156 | | | | 0.156 | | | | | ± 0.020 |
| VGM MCV | µm ³ ; fl | 71 | | | | 72 | | | | | ± 4 |
| TGMH MCH | pg | 25.9 | | | | 25.4 | | | | | ± 2.0 |
| | fmol | 1.61 | | | | 1.58 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.5 | | | | 35.3 | | | | | ± 3.0 |
| | g/l | 365 | | | | 353 | | | | | ± 30 |
| | mmol/l | 22.66 | | | | 21.91 | | | | | ± 1.86 |
| IDR RDW | % | 16.0 | | | | 16.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 66 | | | | 65 | | | | | ± 22 |
| VPM MPV | µm ³ ; fl | 7.9 | | | | 7.5 | | | | | ± 2.0 |
| LYMPHO | # | 1.33 | | | | 1.40 | | | | | ± 0.40 |
| | % | 68.5 | | | | 67.5 | | | | | ± 8.0 |
| MONO | # | 0.20 | | | | 0.20 | | | | | ± 0.20 |
| | % | 8.0 | | | | 8.5 | | | | | ± 6.0 |
| GRANULO | # | 0.47 | | | | 0.50 | | | | | ± 0.40 |
| | % | 23.5 | | | | 24.0 | | | | | ± 7.0 |

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT

MX 445 N
 Rev 1

CONTROL N

(Exp.)

2024-04-05
 (YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|----------------|--------|--------|--------|--------------------------|----------------|--------|--------|--------|-------------------------|
| | | MICROS ES60 | MICROS | MICROS | MICROS | MICROS 60 8P - 60 18P | ARGOS - HELIOS | PENTRA | PENTRA | PENTRA | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 7.7 | | | | 7.6 | | | | | ± 0.8 |
| GR RBC | 10 ⁹ /mm ³ ; 10 ¹² /l | 4.52 | | | | 4.56 | | | | | ± 0.26 |
| HB HGB | g/dl | 13.4 | | | | 13.2 | | | | | ± 0.7 |
| | g/l | 134 | | | | 132 | | | | | ± 7 |
| | mmol/l | 8.32 | | | | 8.22 | | | | | ± 0.43 |
| HT HCT | % | 37.1 | | | | 37.4 | | | | | ± 2.7 |
| | l/l | 0.371 | | | | 0.374 | | | | | ± 0.027 |
| VGM MCV | µm ³ ; fl | 82 | | | | 82 | | | | | ± 4 |
| TGMH MCH | pg | 29.6 | | | | 29.0 | | | | | ± 2.0 |
| | fmol | 1.84 | | | | 1.80 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.2 | | | | 35.4 | | | | | ± 3.0 |
| | g/l | 362 | | | | 354 | | | | | ± 30 |
| | mmol/l | 22.45 | | | | 21.97 | | | | | ± 1.86 |
| IDR RDW | % | 14.5 | | | | 14.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 244 | | | | 260 | | | | | ± 40 |
| VPM MPV | µm ³ ; fl | 7.7 | | | | 7.1 | | | | | ± 2.0 |
| LYMPHO | # | 3.00 | | | | 2.93 | | | | | ± 0.40 |
| | % | 39.0 | | | | 38.5 | | | | | ± 5.0 |
| MONO | # | 0.39 | | | | 0.42 | | | | | ± 0.30 |
| | % | 5.0 | | | | 5.5 | | | | | ± 4.0 |
| GRANULO | # | 4.31 | | | | 4.26 | | | | | ± 0.60 |
| | % | 56.0 | | | | 56.0 | | | | | ± 6.0 |

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

HORIBA ABX SAS
 HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée - BP 7290
 34184 MONTPELLIER Cedex 4
 FRANCE
 Tél.: 33 (0) 4 67 14 15 16
 Fax.: 33 (0) 4 67 14 15 17

LOT

MX 445 H
 Rev 1

CONTROL H

(Exp.)

2024-04-05
 (YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|----------------|--------|--------|--------|--------------------------|----------------|--------|--------|--------|-------------------------|
| | | MICROS ES60 | MICROS | MICROS | MICROS | MICROS 60 8P - 60 18P | ARGOS - HELIOS | PENTRA | PENTRA | PENTRA | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 20.5 | | | | 20.1 | | | | | ± 1.7 |
| GR RBC | 10 ⁹ /mm ³ ; 10 ¹² /l | 5.61 | | | | 5.64 | | | | | ± 0.32 |
| HB HGB | g/dl | 18.5 | | | | 17.9 | | | | | ± 0.9 |
| | g/l | 185 | | | | 179 | | | | | ± 9 |
| | mmol/l | 11.49 | | | | 11.13 | | | | | ± 0.56 |
| HT HCT | % | 50.5 | | | | 50.2 | | | | | ± 3.4 |
| | l/l | 0.505 | | | | 0.502 | | | | | ± 0.034 |
| VGM MCV | µm ³ ; fl | 90 | | | | 89 | | | | | ± 4 |
| TGMH MCH | pg | 33.0 | | | | 31.8 | | | | | ± 2.0 |
| | fmol | 2.05 | | | | 1.97 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.6 | | | | 35.7 | | | | | ± 3.0 |
| | g/l | 366 | | | | 357 | | | | | ± 30 |
| | mmol/l | 22.75 | | | | 22.17 | | | | | ± 1.86 |
| IDR RDW | % | 14.0 | | | | 14.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 492 | | | | 509 | | | | | ± 65 |
| VPM MPV | µm ³ ; fl | 7.9 | | | | 7.1 | | | | | ± 2.0 |
| LYMPHO | # | 3.38 | | | | 3.32 | | | | | ± 1.00 |
| | % | 16.5 | | | | 16.5 | | | | | ± 5.0 |
| MONO | # | 1.74 | | | | 1.61 | | | | | ± 0.80 |
| | % | 8.5 | | | | 8.0 | | | | | ± 4.0 |
| GRANULO | # | 15.38 | | | | 15.18 | | | | | ± 1.40 |
| | % | 75.0 | | | | 75.5 | | | | | ± 7.0 |