

ABX Minotrol CRP



LOT

MC 430

CONTROL



(Exp.) 2024-04-05
(YYYY - MM - DD)

Rev 2

		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1			± 0.4	7.3	7.7			± 1.0	19.8	20.4			± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.21	2.21			± 0.15	4.56	4.52			± 0.18	5.68	5.63			± 0.25
HB HGB	g/dl	5.9	5.9			± 0.4	13.5	13.5			± 0.5	18.0	17.9			± 0.7
	g/l	59	59			± 4	135	135			± 5	180	179			± 7
	mmol/l	3.66	3.66			± 0.25	8.38	8.38			± 0.31	11.18	11.12			± 0.44
HT HCT	%	15.9	15.9			± 2.0	37.4	37.1			± 2.5	51.1	50.7			± 3.0
	l/l	0.159	0.159			± 0.020	0.374	0.371			± 0.025	0.511	0.507			± 0.030
VGM MCV	µm ³ ; fl	72	72			± 4	82	82			± 4	90	90			± 4
TGMH MCH	pg	26.7	26.7			± 2.4	29.6	29.9			± 2.4	31.7	31.8			± 2.8
	fmol	1.66	1.66			± 0.15	1.84	1.85			± 0.15	1.97	1.97			± 0.18
CCMH MCHC	g/dl	37.1	37.1			± 3.0	36.1	36.4			± 3.0	35.2	35.3			± 3.0
	g/l	371	371			± 30	361	364			± 30	352	353			± 30
	mmol/l	23.03	23.03			± 1.86	22.42	22.62			± 1.86	21.87	21.94			± 1.86
IDR RDW	%	13.5	15.0			± 3.0	13.5	13.5			± 3.0	13.5	13.0			± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	72	69			± 20	264	259			± 40	515	511			± 60
VPM MPV	µm ³ ; fl	8.2	7.6			± 2.0	7.4	7.1			± 2.0	7.8	7.5			± 2.0
LYMPHO	#	1.4	1.4			± 0.2	2.7	2.8			± 0.5	3.1	3.2			± 1.1
	%	68.0	67.0			± 8.0	37.0	36.5			± 6.0	15.5	15.5			± 5.0
MONO	#	0.2	0.2			± 0.2	0.6	0.5			± 0.4	1.8	1.8			± 1.0
	%	11.0	9.0			± 7.0	8.0	6.0			± 5.0	9.0	9.0			± 5.0
GRANULO	#	0.4	0.5			± 0.2	4.0	4.4			± 0.5	14.9	15.4			± 1.5
	%	21.0	24.0			± 7.0	55.0	57.5			± 6.0	75.5	75.5			± 7.0
CRP	mg/dl	0.39	N/A			± 0.30	2.19	N/A			± 0.70	4.02	N/A			± 1.00
	mg/l	3.9	N/A			± 3.0	21.9	N/A			± 7.0	40.2	N/A			± 10.0

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170