

ABX Minotrol CRP



LOT

MC 430

CONTROL



(Exp.) 2024-04-05
(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.2		± 0.4	7.3	7.8		± 1.0	19.4	20.5		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.21	2.21		± 0.15	4.56	4.52		± 0.18	5.68	5.57		± 0.25				
HB HGB	g/dl	5.6	5.7		± 0.4	13.2	13.4		± 0.5	18.0	18.1		± 0.7				
	g/l	56	57		± 4	132	134		± 5	180	181		± 7				
	mmol/l	3.48	3.54		± 0.25	8.22	8.32		± 0.31	11.18	11.24		± 0.44				
HT HCT	%	15.9	16.1		± 2.0	37.4	37.5		± 2.5	51.1	50.7		± 3.0				
	l/l	0.159	0.161		± 0.020	0.374	0.375		± 0.025	0.511	0.507		± 0.030				
VGM MCV	µm ³ ; fl	72	73		± 4	82	83		± 4	90	91		± 4				
TGMH MCH	pg	25.4	25.8		± 2.4	29.0	29.6		± 2.4	31.7	32.5		± 2.8				
	fmol	1.57	1.60		± 0.15	1.80	1.84		± 0.15	1.97	2.02		± 0.18				
CCMH MCHC	g/dl	35.2	35.3		± 3.0	35.4	35.7		± 3.0	35.2	35.7		± 3.0				
	g/l	352	353		± 30	354	357		± 30	352	357		± 30				
	mmol/l	21.87	21.94		± 1.86	21.97	22.18		± 1.86	21.87	22.18		± 1.86				
IDR RDW	%	13.5	16.5		± 3.0	13.5	15.0		± 3.0	13.5	14.0		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	72	68		± 20	264	266		± 40	515	515		± 60				
VPM MPV	µm ³ ; fl	8.2	8.6		± 2.0	7.4	7.9		± 2.0	7.8	8.2		± 2.0				
LYMPHO	#	1.5	1.5		± 0.2	2.8	3.0		± 0.5	3.2	3.5		± 1.1				
	%	70.7	68.0		± 8.0	38.9	38.0		± 6.0	16.6	17.0		± 5.0				
MONO	#	0.2	0.2		± 0.2	0.6	0.4		± 0.4	1.7	1.9		± 1.0				
	%	8.8	9.0		± 7.0	7.6	5.5		± 5.0	9.0	9.5		± 5.0				
GRANULO	#	0.4	0.5		± 0.2	3.9	4.4		± 0.5	14.5	15.1		± 1.5				
	%	20.6	23.0		± 7.0	53.9	56.5		± 6.0	74.7	73.5		± 7.0				
CRP	mg/dl	0.39	N/A		± 0.30	2.19	N/A		± 0.70	4.02	N/A		± 1.00				
	mg/l	3.9	N/A		± 3.0	21.9	N/A		± 7.0	40.2	N/A		± 10.0				

1300027267-A
Ref: TEMP-1070 Rev.42