


LOT PX 445

CONTROL

 (Exp.) **2024-03-05**
(YYYY-MM-DD)

Rev 1

| | | ABX Lysebio | | | | | | | | | | | | | | | | | |
|--------------------------|--|-------------|--|----------------------|--|---------|-------------------------|---------|--------------|--|----------------------|---|-------------------------|--------------|--|---------|--|---|-------------------------|
| PARAMETRES PARAMETERS | UNITES UNITS | CONTROL | | | | L | TOLERANCES TOLERANCE | CONTROL | | | | N | TOLERANCES TOLERANCE | CONTROL | | | | H | TOLERANCES TOLERANCE |
| | | PENTRA | | | | | | PENTRA | | | | | | PENTRA | | | | | |
| | | | | DX NEXUS DF NEXUS | | | | | | | DX NEXUS DF NEXUS | | | | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | | | 3.0 | | ± 0.4 | | | 8.9 | | ± 1.0 | | | 19.2 | | ± 2.2 | | | |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | | | 2.37 | | ± 0.16 | | | 4.62 | | ± 0.20 | | | 5.08 | | ± 0.25 | | | |
| HB HGB | g/dl | | | 6.2 | | ± 0.4 | | | 13.6 | | ± 0.5 | | | 16.0 | | ± 0.6 | | | |
| | g/l | | | 62 | | ± 4 | | | 136 | | ± 5 | | | 160 | | ± 6 | | | |
| | mmol/l | | | 3.85 | | ± 0.25 | | | 8.45 | | ± 0.31 | | | 9.94 | | ± 0.37 | | | |
| HT HCT | % | | | 18.7 | | ± 1.5 | | | 39.7 | | ± 2.0 | | | 46.7 | | ± 2.5 | | | |
| | l/l | | | 0.187 | | ± 0.015 | | | 0.397 | | ± 0.020 | | | 0.467 | | ± 0.025 | | | |
| VGM MCV | µm ³ ; fl | | | 79 | | ± 5 | | | 86 | | ± 5 | | | 92 | | ± 5 | | | |
| TGMH MCH | pg | | | 26.2 | | ± 2.0 | | | 29.4 | | ± 2.0 | | | 31.5 | | ± 2.5 | | | |
| | fmol | | | 1.62 | | ± 0.12 | | | 1.83 | | ± 0.12 | | | 1.96 | | ± 0.16 | | | |
| CCMH MCHC | g/dl | | | 33.1 | | ± 3.0 | | | 34.2 | | ± 3.0 | | | 34.2 | | ± 3.0 | | | |
| | g/l | | | 331 | | ± 30 | | | 342 | | ± 30 | | | 342 | | ± 30 | | | |
| | mmol/l | | | 20.56 | | ± 1.86 | | | 21.26 | | ± 1.86 | | | 21.26 | | ± 1.86 | | | |
| IDR RDW | % | | | 18.5 | | ± 4.0 | | | 15.5 | | ± 4.0 | | | 13.5 | | ± 4.0 | | | |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | | | 72 | | ± 20 | | | 237 | | ± 30 | | | 463 | | ± 50 | | | |
| VPM MPV | µm ³ ; fl | | | 9.8 | | ± 2.0 | | | 9.2 | | ± 2.0 | | | 9.8 | | ± 2.0 | | | |
| NEUT | # | | | 1.58 | | ± 0.35 | | | 4.53 | | ± 0.90 | | | 13.50 | | ± 1.90 | | | |
| | % | | | 51.1 | | ± 10.0 | | | 50.9 | | ± 10.0 | | | 70.0 | | ± 10.0 | | | |
| LYMPHO | # | | | 0.98 | | ± 0.33 | | | 3.36 | | ± 0.70 | | | 3.53 | | ± 1.50 | | | |
| | % | | | 31.7 | | ± 12.0 | | | 37.8 | | ± 8.0 | | | 18.4 | | ± 8.0 | | | |
| MONO | # | | | 0.18 | | ± 0.18 | | | 0.33 | | ± 0.33 | | | 0.44 | | ± 0.44 | | | |
| | % | | | 5.7 | | ± 5.7 | | | 3.7 | | ± 3.7 | | | 2.3 | | ± 2.3 | | | |
| EOS | # | | | 0.21 | | ± 0.21 | | | 0.28 | | ± 0.28 | | | 0.84 | | ± 0.84 | | | |
| | % | | | 6.8 | | ± 6.8 | | | 3.2 | | ± 3.2 | | | 4.4 | | ± 4.4 | | | |
| BASO | # | | | 0.15 | | ± 0.15 | | | 0.39 | | ± 0.39 | | | 0.94 | | ± 0.94 | | | |
| | % | | | 4.7 | | ± 4.7 | | | 4.4 | | ± 4.4 | | | 4.9 | | ± 4.9 | | | |

Ref: TEMP-0821 Rev.50 BACK / VERSO