

ABX Minocal



LOT		CX 490		CAL		(Exp.) 2024-02-05 (YYYY - MM - DD)						
		Rev 1		ABX Lysebio								
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60		60 60 C+ ES60	XL80		
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.2	9.3	9.2	9.2	9.3	9.1		9.2	9.2	± 0.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.43	4.46	4.40	4.45	4.46	4.39		4.48	4.45	± 0.06	
HB HGB	g/dl	13.0	13.0	12.9	13.1	13.0	12.9		13.1	13.1	± 0.2	
	g/l	130	130	129	131	130	129		131	131	± 2	
	mmol/l	8.07	8.07	8.01	8.14	8.07	8.01		8.14	8.14	± 0.12	
HT HCT	%	36.5	36.8	35.2	36.0	35.7	35.1		35.8	36.0	± 1.0	
	l/l	0.365	0.368	0.352	0.360	0.357	0.351		0.358	0.360	± 0.010	
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	238	244	234	236	243	231		236	236	± 10	
VMP MPV	µm ³ ; fl	8.5	8.9	N/A	N/A	N/A	N/A		N/A	N/A	± 0.5	

1300090308
BACK / VERSO
Ref. TEMP-0387 Rev.51