

# ABX Minotrol Retic



**LOT** RX 445  
Rev 1

**CONTROL**

(Exp.) 2024-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
				1				2				3	
				PENTRA				PENTRA				PENTRA	
		DX Nexus			DX Nexus			DX Nexus					
Méthode Manuelle / Manual RETIC	%			1.4	± 0.7			5.4	± 2.2			10.4	± 3.2
RETIC	%			1.37	± 0.69			4.78	± 1.20			9.55	± 1.91
RETIC	10 <sup>6</sup> /µl			0.048	± 0.024			0.167	± 0.042			0.332	± 0.066
MFI MRI *	%			25.6	± 8.0			30.7	± 10.0			32.3	± 12.0
GR RBC	10 <sup>6</sup> /µl			3.49	± 0.30			3.50	± 0.30			3.48	± 0.30
PIC RANGE	-			27.0	± 5.0			27.0	± 5.0			27.0	± 5.0

\* For research use only in Canada and USA

Ref: TEMP-0382 Rev.45 FRONT / RECTO 1300090305

# ABX Minotrol Retic



**LOT** RX 445  
Rev 1

**CONTROL**

(Exp.) 2024-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITS UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	5.4			± 2.2	10.4			± 3.2
RETIC	%	1.52			± 0.76	5.09			± 1.27	10.18			± 2.04
RETIC	10 <sup>6</sup> /µl	0.052			± 0.026	0.174			± 0.044	0.344			± 0.069
MFI MRI *	%	22.2			± 8.0	28.1			± 10.0	29.3			± 12.0
GR RBC	10 <sup>6</sup> /µl	3.40			± 0.30	3.41			± 0.30	3.38			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.45 BACK / VERSO 1300090305

\* For research use only in Canada and USA