


LOT PX 445
Rev 1


CONTROL

 (Exp.) 2024-03-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0			± 0.4	8.7	8.7			± 1.0	18.6	18.6			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.30	2.30			± 0.16	4.54	4.54			± 0.20	4.93	4.93			± 0.25			
HB HGB	g/dl	6.0	6.0			± 0.4	13.5	13.5			± 0.5	16.0	16.0			± 0.6			
	g/l	60	60			± 4	135	135			± 5	160	160			± 6			
	mmol/l	3.73	3.73			± 0.25	8.38	8.38			± 0.31	9.94	9.94			± 0.37			
HT HCT	%	17.9	17.9			± 1.5	38.6	38.6			± 2.0	44.9	44.9			± 2.5			
	l/l	0.179	0.179			± 0.015	0.386	0.386			± 0.020	0.449	0.449			± 0.025			
VGM MCV	µm ³ ; fl	78	78			± 5	85	85			± 5	91	91			± 5			
TGMH MCH	pg	26.1	26.1			± 2.0	29.7	29.7			± 2.0	32.5	32.5			± 2.5			
	fmol	1.62	1.62			± 0.12	1.85	1.85			± 0.12	2.02	2.02			± 0.16			
CCMH MCHC	g/dl	33.4	33.4			± 3.0	35.0	35.0			± 3.0	35.7	35.7			± 3.0			
	g/l	334	334			± 30	350	350			± 30	357	357			± 30			
	mmol/l	20.77	20.77			± 1.86	21.72	21.72			± 1.86	22.15	22.15			± 1.86			
IDR RDW	%	15.5	15.5			± 4.0	13.5	13.5			± 4.0	12.0	12.0			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	69	69			± 20	230	230			± 30	441	441			± 50			
VPM MPV	µm ³ ; fl	9.7	9.7			± 2.0	9.5	9.5			± 2.0	10.2	10.2			± 2.0			
NEUT	#	1.46	1.46			± 0.35	4.38	4.38			± 0.90	12.95	12.95			± 1.90			
	%	48.6	48.6			± 10.0	50.3	50.3			± 10.0	69.6	69.6			± 10.0			
LYMPHO	#	1.15	1.15			± 0.33	3.53	3.53			± 0.70	3.70	3.70			± 1.50			
	%	38.4	38.4			± 12.0	40.6	40.6			± 8.0	19.9	19.9			± 8.0			
MONO	#	0.15	0.15			± 0.15	0.28	0.28			± 0.28	0.37	0.37			± 0.37			
	%	4.9	4.9			± 4.9	3.2	3.2			± 3.2	2.0	2.0			± 2.0			
EOS	#	0.16	0.16			± 0.16	0.27	0.27			± 0.27	0.84	0.84			± 0.84			
	%	5.4	5.4			± 5.4	3.1	3.1			± 3.1	4.5	4.5			± 4.5			
BASO	#	0.08	0.08			± 0.08	0.24	0.24			± 0.24	0.74	0.74			± 0.74			
	%	2.7	2.7			± 2.7	2.8	2.8			± 2.8	4.0	4.0			± 4.0			

LOT PX 445
Rev 1

CONTROL

 (Exp.) 2024-03-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCRP		60 60C+ ES60		MS60	MSCRP		60 60C+ ES60	MS60		MSCRP		60 60C+ ES60	MS60	MSCRP	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0	3.0		± 0.4	8.8	8.7	8.7		± 1.0	19.0	18.8	18.8		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.33	2.35		± 0.16	4.57	4.51	4.53		± 0.20	4.96	4.89	4.95		± 0.25			
	g/dl	6.1	6.0	6.1		± 0.4	13.6	13.4	13.6		± 0.5	16.1	15.8	16.1		± 0.6			
HB HGB	g/l	61	60	61		± 4	136	134	136		± 5	161	158	161		± 6			
	mmol/l	3.79	3.73	3.79		± 0.25	8.45	8.32	8.45		± 0.31	10.00	9.81	10.00		± 0.37			
HT HCT	%	17.9	17.5	17.6		± 1.5	38.8	38.3	38.5		± 2.0	46.1	45.0	45.3		± 2.5			
	l/l	0.179	0.175	0.176		± 0.015	0.388	0.383	0.385		± 0.020	0.461	0.450	0.453		± 0.025			
VGM MCV	µm ³ ; fl	75	75	75.0		± 5	85	85	85.0		± 5	93	92	91.5		± 5			
TGMH MCH	pg	25.5	25.8	26.0		± 2.0	29.8	29.7	30.0		± 2.0	32.5	32.3	32.5		± 2.5			
	fmol	1.58	1.60	1.61		± 0.12	1.85	1.85	1.86		± 0.12	2.02	2.01	2.02		± 0.16			
	g/dl	34.0	34.3	34.6		± 3.0	35.0	35.0	35.3		± 3.0	34.9	35.1	35.5		± 3.0			
CCMH MCHC	g/l	340	343	346		± 30	350	350	353		± 30	349	351	355		± 30			
	mmol/l	21.13	21.32	21.49		± 1.86	21.74	21.71	21.93		± 1.86	21.67	21.81	22.07		± 1.86			
IDR RDW	%	16.0	16.0	15.0		± 4.0	13.0	13.5	12.5		± 4.0	12.0	12.0	11.5		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	71	68		± 20	234	229	218		± 30	460	456	429		± 50			
VPM MPV	µm ³ ; fl	9.5	9.4	8.3		± 2.0	8.7	8.9	8.2		± 2.0	9.3	9.3	8.5		± 2.0			
NEUT	#	1.45	1.43	1.43		± 0.35	4.35	4.33	4.31		± 0.90	13.07	13.10	13.16		± 1.90			
	%	48.3	47.6	46.0		± 10.0	49.4	49.8	49.5		± 10.0	68.8	69.7	70.0		± 10.0			
LYMPHO	#	1.17	1.17	1.24		± 0.33	3.64	3.60	3.57		± 0.70	3.88	3.85	3.76		± 1.50			
	%	38.9	39.1	40.0		± 12.0	41.4	41.4	41.0		± 8.0	20.4	20.5	20.0		± 8.0			
MONO	#	0.15	0.14	0.17		± 0.14	0.27	0.26	0.26		± 0.26	0.40	0.32	0.28		± 0.28			
	%	4.9	4.8	5.5		± 4.8	3.1	3.0	3.0		± 3.0	2.1	1.7	1.5		± 1.5			
EOS	#	0.16	0.17	0.17		± 0.16	0.29	0.26	0.30		± 0.26	0.89	0.79	0.85		± 0.79			
	%	5.2	5.8	5.5		± 5.2	3.3	3.0	3.5		± 3.0	4.7	4.2	4.5		± 4.2			
BASO	#	0.08	0.08	0.09		± 0.08	0.25	0.24	0.26		± 0.24	0.79	0.73	0.75		± 0.73			
	%	2.7	2.7	3.0		± 2.7	2.8	2.8	3.0		± 2.8	4.0	3.9	4.0		± 3.9			

Ref: TEMP-0821 Rev.50 FRONT / RECTO