


LOT PX 442
Rev 1

CONTROL

 (Exp.) 2023-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.40	2.40		± 0.40	7.20		7.20	± 1.00		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.46	2.46	± 0.16	4.72	4.72	± 0.20	5.32	5.32	± 0.25	
HB HGB	g/dl	6.2	6.2	± 0.4	14.0	14.0	± 0.5	17.0	17.0	± 0.6	
	g/l	62	62	± 4	140	140	± 5	170	170	± 6	
	mmol/l	3.85	3.85	± 0.25	8.69	8.69	± 0.31	10.56	10.56	± 0.37	
HT HCT	%	18.6	18.6	± 1.5	41.1	41.1	± 2.0	49.7	49.7	± 2.5	
	l/l	0.186	0.186	± 0.015	0.411	0.411	± 0.020	0.497	0.497	± 0.025	
VGM MCV	µm ³ ; fl	75.5	75.5	± 5.0	87.0	87.0	± 5.0	93.5	93.5	± 5.0	
TGMH MCH	pg	25.2	25.2	± 2.0	29.7	29.7	± 2.0	32.0	32.0	± 2.5	
	fmol	1.57	1.57	± 0.12	1.84	1.84	± 0.12	1.98	1.98	± 0.16	
CCMH MCHC	g/dl	33.4	33.4	± 3.0	34.1	34.1	± 3.0	34.2	34.2	± 3.0	
	g/l	334	334	± 30	341	341	± 30	342	342	± 30	
	mmol/l	20.73	20.73	± 1.86	21.17	21.17	± 1.86	21.22	21.22	± 1.86	
IDR-CV RDW-CV	%	14.5	14.5	± 4.0	11.5	11.5	± 4.0	11.5	11.5	± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	67	67	± 20	258	258	± 30	482	482	± 50	
VPM MPV	µm ³ ; fl	8.7	8.7	± 2.0	9.3	9.3	± 2.0	9.4	9.4	± 2.0	
NEUT	#	1.29	1.29	± 0.35	4.06	4.06	± 0.90	12.76	12.76	± 1.90	
	%	53.8	53.8	± 10.0	56.4	56.4	± 10.0	74.2	74.2	± 10.0	
LYMPHO	#	0.67	0.67	± 0.33	2.14	2.14	± 0.70	2.08	2.08	± 1.50	
	%	27.9	27.9	± 12.0	29.7	29.7	± 8.0	12.1	12.1	± 8.0	
MONO	#	0.17	0.17	± 0.17	0.48	0.48	± 0.43	0.76	0.76	± 0.68	
	%	7.1	7.1	± 7.1	6.6	6.6	± 5.9	4.4	4.4	± 4.0	
EOS	#	0.17	0.17	± 0.17	0.25	0.25	± 0.25	0.83	0.83	± 0.83	
	%	7.2	7.2	± 7.2	3.5	3.5	± 3.5	4.8	4.8	± 4.8	
BASO	#	0.10	0.10	± 0.10	0.27	0.27	± 0.18	0.77	0.77	± 0.39	
	%	4.0	4.0	± 2.5	3.8	3.8	± 3.0	4.5	4.5	± 2.5	
IDR-SD RDW-SD	fl	43.0	43.0	± 8.0	41.0	41.0	± 8.0	42.5	42.5	± 8.0	
ERB NRBC	#	0.38	0.38	± 0.30	1.25	1.25	± 0.45	0.92	0.92	± 0.35	
	%	16.0	16.0	± 8.0	17.4	17.4	± 8.0	5.4	5.4	± 2.0	
PLTS OPT	10 ³ /mm ³ ; 10 ⁹ /l	N/A	67	± 30	N/A	252	± 40	N/A	474	± 50	

BACK / VERSO
Ref: TEMP-0821 Rev.48