


LOT PX 443
Rev 2

CONTROL

 (Exp.) **2023-11-05**
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500 Since V1.7	H2500 Since V1.7		H1500 Since V1.7	H2500 Since V1.7		H1500 Since V1.7	H2500 Since V1.7		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.57	2.57		± 0.40	7.59		7.59	± 1.00		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.26	2.26	± 0.16	4.64	4.64	± 0.20	5.30	5.30	± 0.25	
HB HGB	g/dl	5.9	5.9	± 0.4	13.9	13.9	± 0.5	17.0	17.0	± 0.6	
	g/l	59	59	± 4	139	139	± 5	170	170	± 6	
	mmol/l	3.66	3.66	± 0.25	8.63	8.63	± 0.31	10.56	10.56	± 0.37	
HT HCT	%	17.6	17.6	± 1.5	40.5	40.5	± 2.0	50.6	50.6	± 2.5	
	l/l	0.176	0.176	± 0.015	0.405	0.405	± 0.020	0.506	0.506	± 0.025	
VGM MCV	µm ³ ; fl	77.7	77.7	± 5.0	87.4	87.4	± 5.0	95.5	95.5	± 5.0	
TGMH MCH	pg	26.1	26.1	± 2.0	30.0	30.0	± 2.0	32.1	32.1	± 2.5	
	fmol	1.62	1.62	± 0.12	1.86	1.86	± 0.12	1.99	1.99	± 0.16	
CCMH MCHC	g/dl	33.6	33.6	± 3.0	34.3	34.3	± 3.0	33.6	33.6	± 3.0	
	g/l	336	336	± 30	343	343	± 30	336	336	± 30	
	mmol/l	20.86	20.86	± 1.86	21.29	21.29	± 1.86	20.86	20.86	± 1.86	
IDR-SD RDW-SD	fl	40.3	40.3	± 8.0	37.1	37.1	± 8.0	38.5	38.5	± 8.0	
IDR-CV RDW-CV	%	16.5	16.5	± 4.0	13.2	13.2	± 4.0	12.2	12.2	± 4.0	
PLA PLT	10 ⁹ /mm ³ ; 10 ⁹ /l	66	66	± 20	262	262	± 30	514	514	± 50	
PLT-Ox	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	66	± 30	N/A	253	± 40	N/A	490	± 50	
VMP MPV	µm ³ ; fl	10.6	10.6	± 2.0	10.3	10.3	± 2.0	11.1	11.1	± 2.0	
LPF	%	N/A	1.8	± 1.8	N/A	1.7	± 1.7	N/A	2.2	± 2.2	
NEU	10 ⁹ /mm ³ ; 10 ⁹ /l	1.46	1.46	± 0.35	4.46	4.46	± 0.90	12.83	12.83	± 1.90	
	%	56.7	56.7	± 10.0	58.8	58.8	± 10.0	71.4	71.4	± 10.0	
LYM	10 ⁹ /mm ³ ; 10 ⁹ /l	0.56	0.56	± 0.33	1.70	1.70	± 0.70	1.78	1.78	± 1.50	
	%	21.7	21.7	± 12.0	22.4	22.4	± 8.0	9.9	9.9	± 8.0	
MON	10 ⁹ /mm ³ ; 10 ⁹ /l	0.24	0.24	± 0.24	0.60	0.60	± 0.54	1.12	1.12	± 1.01	
	%	9.3	9.3	± 9.30	7.9	7.9	± 7.1	6.2	6.2	± 5.6	
EOS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.14	0.14	± 0.14	0.28	0.28	± 0.28	0.72	0.72	± 0.72	
	%	5.4	5.4	± 5.40	3.7	3.7	± 3.7	4.0	4.0	± 4.0	
BAS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.11	0.11	± 0.11	0.33	0.33	± 0.22	0.83	0.83	± 0.42	
	%	4.1	4.1	± 2.5	4.4	4.4	± 3.0	4.6	4.6	± 2.5	
IMG	10 ⁹ /mm ³ ; 10 ⁹ /l	0.07	0.07	± 0.07	0.22	0.22	± 0.20	0.71	0.71	± 0.64	
	%	2.8	2.8	± 2.80	2.9	2.9	± 2.6	3.9	3.9	± 3.5	
ERB NRBC	10 ⁹ /mm ³ ; 10 ⁹ /l	0.24	0.24	± 0.24	0.73	0.73	± 0.45	0.76	0.76	± 0.35	
	%	9.3	9.3	± 8.0	9.6	9.6	± 8.0	4.2	4.2	± 2.0	
TNC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.81	2.81	± 0.40	8.32	8.32	± 1.00	18.74	18.74	± 2.20	

Ref: TEMP-0821 Rev.49