

# ABX Difftrol



**LOT** PX 436  
Rev 1

**CONTROL**

(Exp.) **2022-09-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio														
		CONTROL		L	TOLERANCES TOLERANCE	CONTROL		N	TOLERANCES TOLERANCE	CONTROL		H	TOLERANCES TOLERANCE			
		YUMIZEN		H1500		H2500	YUMIZEN			H1500	H2500	YUMIZEN		H1500	H2500	
		H1500	H2500				H1500	H2500				H1500				H2500
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.40</b>	<b>2.40</b>		± 0.40	<b>7.40</b>	<b>7.40</b>		± 1.00	<b>17.80</b>	<b>17.80</b>		± 2.20			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.44</b>	<b>2.44</b>		± 0.16	<b>4.76</b>	<b>4.76</b>		± 0.20	<b>5.24</b>	<b>5.24</b>		± 0.25			
HB HGB	g/dl	<b>6.3</b>	<b>6.3</b>		± 0.4	<b>13.7</b>	<b>13.7</b>		± 0.5	<b>16.3</b>	<b>16.3</b>		± 0.6			
	g/l	<b>63</b>	<b>63</b>		± 4	<b>137</b>	<b>137</b>		± 5	<b>163</b>	<b>163</b>		± 6			
	mmol/l	<b>3.91</b>	<b>3.91</b>		± 0.25	<b>8.51</b>	<b>8.51</b>		± 0.31	<b>10.12</b>	<b>10.12</b>		± 0.37			
HT HCT	%	<b>19.2</b>	<b>19.2</b>		± 1.5	<b>41.1</b>	<b>41.1</b>		± 2.0	<b>48.8</b>	<b>48.8</b>		± 2.5			
	l/l	<b>0.192</b>	<b>0.192</b>		± 0.015	<b>0.411</b>	<b>0.411</b>		± 0.020	<b>0.488</b>	<b>0.488</b>		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	<b>78.6</b>	<b>78.6</b>		± 5.0	<b>86.6</b>	<b>86.6</b>		± 5.0	<b>93.5</b>	<b>93.5</b>		± 5.0			
TGMH MCH	pg	<b>25.8</b>	<b>25.8</b>		± 2.0	<b>28.8</b>	<b>28.8</b>		± 2.0	<b>31.1</b>	<b>31.1</b>		± 2.5			
	fmol	<b>1.60</b>	<b>1.60</b>		± 0.12	<b>1.79</b>	<b>1.79</b>		± 0.12	<b>1.93</b>	<b>1.93</b>		± 0.16			
CCMH MCHC	g/dl	<b>32.8</b>	<b>32.8</b>		± 3.0	<b>33.2</b>	<b>33.2</b>		± 3.0	<b>33.3</b>	<b>33.3</b>		± 3.0			
	g/l	<b>328</b>	<b>328</b>		± 30	<b>332</b>	<b>332</b>		± 30	<b>333</b>	<b>333</b>		± 30			
	mmol/l	<b>20.40</b>	<b>20.40</b>		± 1.86	<b>20.64</b>	<b>20.64</b>		± 1.86	<b>20.66</b>	<b>20.66</b>		± 1.86			
IDR-CV RDW-CV	%	<b>14.5</b>	<b>14.5</b>		± 4.0	<b>12.5</b>	<b>12.5</b>		± 4.0	<b>11.5</b>	<b>11.5</b>		± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>71</b>	<b>71</b>		± 20	<b>257</b>	<b>257</b>		± 30	<b>512</b>	<b>512</b>		± 50			
VPM MPV	µm <sup>3</sup> ; fl	<b>9.7</b>	<b>9.7</b>		± 2.0	<b>12.7</b>	<b>12.7</b>		± 2.0	<b>11.6</b>	<b>11.6</b>		± 2.0			
NEUT	#	<b>1.21</b>	<b>1.21</b>		± 0.35	<b>3.94</b>	<b>3.94</b>		± 0.90	<b>12.87</b>	<b>12.87</b>		± 1.90			
	%	<b>50.6</b>	<b>50.6</b>		± 10.0	<b>53.3</b>	<b>53.3</b>		± 10.0	<b>72.3</b>	<b>72.3</b>		± 10.0			
LYMPHO	#	<b>0.68</b>	<b>0.68</b>		± 0.33	<b>2.39</b>	<b>2.39</b>		± 0.70	<b>2.47</b>	<b>2.47</b>		± 1.50			
	%	<b>28.3</b>	<b>28.3</b>		± 12.0	<b>32.3</b>	<b>32.3</b>		± 8.0	<b>13.9</b>	<b>13.9</b>		± 8.0			
MONO	#	<b>0.23</b>	<b>0.23</b>		± 0.23	<b>0.44</b>	<b>0.44</b>		± 0.40	<b>0.73</b>	<b>0.73</b>		± 0.66			
	%	<b>9.6</b>	<b>9.6</b>		± 9.6	<b>6.0</b>	<b>6.0</b>		± 5.4	<b>4.1</b>	<b>4.1</b>		± 3.7			
EOS	#	<b>0.18</b>	<b>0.18</b>		± 0.18	<b>0.33</b>	<b>0.33</b>		± 0.33	<b>0.94</b>	<b>0.94</b>		± 0.94			
	%	<b>7.4</b>	<b>7.4</b>		± 7.4	<b>4.4</b>	<b>4.4</b>		± 4.4	<b>5.3</b>	<b>5.3</b>		± 5.3			
BASO	#	<b>0.10</b>	<b>0.10</b>		± 0.10	<b>0.30</b>	<b>0.30</b>		± 0.20	<b>0.78</b>	<b>0.78</b>		± 0.39			
	%	<b>4.1</b>	<b>4.1</b>		± 2.5	<b>4.0</b>	<b>4.0</b>		± 3.0	<b>4.4</b>	<b>4.4</b>		± 2.5			
IDR-SD RDW-SD	fl	<b>43.0</b>	<b>43.0</b>		± 4.0	<b>41.0</b>	<b>41.0</b>		± 4.0	<b>41.5</b>	<b>41.5</b>		± 4.0			
ERB NRBC	#	<b>0.42</b>	<b>0.42</b>		± 0.13	<b>1.48</b>	<b>1.48</b>		± 0.30	<b>1.12</b>	<b>1.12</b>		± 0.25			
	%	<b>17.7</b>	<b>17.7</b>		± 4.0	<b>20.0</b>	<b>20.0</b>		± 5.0	<b>6.3</b>	<b>6.3</b>		± 1.5			
PLTS OPT	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	N/A	<b>70</b>		± 30	N/A	<b>220</b>		± 40	N/A	<b>451</b>		± 50			

1300105209 BACK / VERSO Ref: TEMP-0821 Rev.46