


LOT PX 438
Rev 1

CONTROL

 (Exp.) 2023-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500	H2500		H1500	H2500		H1500	H2500		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.50	2.50		± 0.40	7.40		7.40	± 1.00		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.41	± 0.16	4.68	4.68	± 0.20	5.19	5.19	± 0.25	
HB HGB	g/dl	6.3	6.3	± 0.4	14.3	14.3	± 0.5	16.5	16.5	± 0.6	
	g/l	63	63	± 4	143	143	± 5	165	165	± 6	
	mmol/l	3.91	3.91	± 0.25	8.88	8.88	± 0.31	10.25	10.25	± 0.37	
HT HCT	%	18.9	18.9	± 1.5	42.1	42.1	± 2.0	48.3	48.3	± 2.5	
	l/l	0.189	0.189	± 0.015	0.421	0.421	± 0.020	0.483	0.483	± 0.025	
VGM MCV	µm ³ ; fl	78.5	78.5	± 5.0	90.0	90.0	± 5.0	93.0	93.0	± 5.0	
TGMH MCH	pg	26.1	26.1	± 2.0	30.6	30.6	± 2.0	31.8	31.8	± 2.5	
	fmol	1.62	1.62	± 0.12	1.90	1.90	± 0.12	1.97	1.97	± 0.16	
CCMH MCHC	g/dl	33.3	33.3	± 3.0	34.0	34.0	± 3.0	34.2	34.2	± 3.0	
	g/l	333	333	± 30	340	340	± 30	342	342	± 30	
	mmol/l	20.68	20.68	± 1.86	21.08	21.08	± 1.86	21.23	21.23	± 1.86	
IDR-CV RDW-CV	%	13.5	13.5	± 4.0	12.0	12.0	± 4.0	10.5	10.5	± 4.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	68	± 20	240	240	± 30	490	490	± 50	
VPM MPV	µm ³ ; fl	9.1	9.1	± 2.0	9.3	9.3	± 2.0	8.3	8.3	± 2.0	
NEUT	#	1.30	1.30	± 0.35	4.05	4.05	± 0.90	12.40	12.40	± 1.90	
	%	51.9	51.9	± 10.0	54.7	54.7	± 10.0	72.5	72.5	± 10.0	
LYMPHO	#	0.71	0.71	± 0.33	2.29	2.29	± 0.70	2.51	2.51	± 1.50	
	%	28.5	28.5	± 12.0	30.9	30.9	± 8.0	14.7	14.7	± 8.0	
MONO	#	0.20	0.20	± 0.20	0.49	0.49	± 0.44	0.68	0.68	± 0.61	
	%	8.0	8.0	± 8.0	6.6	6.6	± 5.9	4.0	4.0	± 3.6	
EOS	#	0.19	0.19	± 0.19	0.29	0.29	± 0.29	0.77	0.77	± 0.77	
	%	7.4	7.4	± 7.4	3.9	3.9	± 3.9	4.5	4.5	± 4.5	
BASO	#	0.11	0.11	± 0.11	0.29	0.29	± 0.19	0.74	0.74	± 0.37	
	%	4.2	4.2	± 2.5	3.9	3.9	± 3.0	4.3	4.3	± 2.5	
IDR-SD RDW-SD	fl	42.5	42.5	± 4.0	41.0	41.0	± 4.0	40.5	40.5	± 4.0	
ERB NRBC	#	0.41	0.41	± 0.13	1.35	1.35	± 0.30	1.13	1.13	± 0.25	
	%	16.3	16.3	± 4.0	18.3	18.3	± 5.0	6.6	6.6	± 1.5	
PLTS OPT	10 ³ /mm ³ ; 10 ⁹ /l	N/A	67	± 30	N/A	224	± 40	N/A	468	± 50	

1300105209
BACK / VERSO
Ref: TEMP-0821 Rev.47