


LOT PX 441
Rev 2

CONTROL

 (Exp.) 2023-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500	H2500		H1500	H2500		H1500	H2500		
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.35	2.35	± 0.40	6.90	6.90	± 1.00	17.50	17.50	± 2.20	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40	± 0.16	4.60	4.60	± 0.20	5.11	5.11	± 0.25	
HB	HGB	g/dl	6.1	6.1	± 0.4	13.7	13.7	± 0.5	16.4	16.4	± 0.6	
		g/l	61	61	± 4	137	137	± 5	164	164	± 6	
		mmol/l	3.79	3.79	± 0.25	8.51	8.51	± 0.31	10.18	10.18	± 0.37	
HT	HCT	%	18.2	18.2	± 1.5	40.5	40.5	± 2.0	48.3	48.3	± 2.5	
		l/l	0.182	0.182	± 0.015	0.405	0.405	± 0.020	0.483	0.483	± 0.025	
VGM	MCV	µm ³ ; fl	76.0	76.0	± 5.0	88.0	88.0	± 5.0	94.5	94.5	± 5.0	
TGMH	MCH	pg	25.4	25.4	± 2.0	29.8	29.8	± 2.0	32.1	32.1	± 2.5	
		fmol	1.58	1.58	± 0.12	1.85	1.85	± 0.12	1.99	1.99	± 0.16	
CCMH	MCHC	g/dl	33.4	33.4	± 3.0	33.8	33.8	± 3.0	34.0	34.0	± 3.0	
		g/l	334	334	± 30	338	338	± 30	340	340	± 30	
		mmol/l	20.77	20.77	± 1.86	21.02	21.02	± 1.86	21.09	21.09	± 1.86	
IDR-CV	RDW-CV	%	14.0	14.0	± 4.0	12.0	12.0	± 4.0	11.0	11.0	± 4.0	
PLAQ.	PLTS	10 ³ /mm ³ ; 10 ⁹ /l	73	73	± 20	244	244	± 30	495	495	± 50	
VPM	MPV	µm ³ ; fl	8.7	8.7	± 2.0	8.9	8.9	± 2.0	8.9	8.9	± 2.0	
NEUT		#	1.13	1.13	± 0.35	3.74	3.74	± 0.90	12.58	12.58	± 1.90	
		%	48.2	48.2	± 10.0	54.2	54.2	± 10.0	71.9	71.9	± 10.0	
LYMPHO		#	0.77	0.77	± 0.33	2.19	2.19	± 0.70	2.35	2.35	± 1.50	
		%	32.6	32.6	± 12.0	31.7	31.7	± 8.0	13.4	13.4	± 8.0	
MONO		#	0.19	0.19	± 0.19	0.43	0.43	± 0.39	0.96	0.96	± 0.86	
		%	8.0	8.0	± 8.0	6.3	6.3	± 5.7	5.5	5.5	± 5.0	
EOS		#	0.17	0.17	± 0.17	0.26	0.26	± 0.26	0.84	0.84	± 0.84	
		%	7.1	7.1	± 7.1	3.8	3.8	± 3.8	4.8	4.8	± 4.8	
BASO		#	0.10	0.10	± 0.10	0.28	0.28	± 0.19	0.77	0.77	± 0.39	
		%	4.1	4.1	± 2.5	4.0	4.0	± 3.0	4.4	4.4	± 2.5	
IDR-SD	RDW-SD	fl	41.0	41.0	± 4.0	41.0	41.0	± 4.0	41.0	41.0	± 4.0	
ERB	NRBC	#	0.46	0.46	± 0.13	1.33	1.33	± 0.30	1.02	1.02	± 0.25	
		%	19.4	19.4	± 4.0	19.3	19.3	± 5.0	5.8	5.8	± 1.5	
PLTS	OPT	10 ³ /mm ³ ; 10 ⁹ /l	N/A	82	± 30	N/A	247	± 40	N/A	499	± 50	

1300105209
BACK / VERSO
Ref: TEMP-0821 Rev.47