


**LOT** PX 444  
Rev 1

**CONTROL**

 (Exp.) 2024-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500 V1.0 to V1.4	H2500 V1.0 to V1.4		H1500 V1.0 to V1.4	H2500 V1.0 to V1.4		H1500 V1.0 to V1.4	H2500 V1.0 to V1.4		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>2.45</b>	<b>2.45</b>	± 0.40		<b>7.30</b>	<b>7.30</b>		± 1.00	<b>17.80</b>		
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.39</b>	<b>2.39</b>	± 0.16	<b>4.64</b>	<b>4.64</b>	± 0.20	<b>5.16</b>	<b>5.16</b>	± 0.25		
HB HGB	g/dl	<b>6.3</b>	<b>6.3</b>	± 0.4	<b>13.6</b>	<b>13.6</b>	± 0.5	<b>16.2</b>	<b>16.2</b>	± 0.6		
	g/l	<b>63</b>	<b>63</b>	± 4	<b>136</b>	<b>136</b>	± 5	<b>162</b>	<b>162</b>	± 6		
HT HCT	mmol/l	<b>3.91</b>	<b>3.91</b>	± 0.25	<b>8.45</b>	<b>8.45</b>	± 0.31	<b>10.06</b>	<b>10.06</b>	± 0.37		
	%	<b>18.8</b>	<b>18.8</b>	± 1.5	<b>39.7</b>	<b>39.7</b>	± 2.0	<b>48.0</b>	<b>48.0</b>	± 2.5		
VGM MCV	l/l	<b>0.188</b>	<b>0.188</b>	± 0.015	<b>0.397</b>	<b>0.397</b>	± 0.020	<b>0.480</b>	<b>0.480</b>	± 0.025		
	µm <sup>3</sup> ; fl	<b>78.5</b>	<b>78.5</b>	± 5.0	<b>85.5</b>	<b>85.5</b>	± 5.0	<b>93.0</b>	<b>93.0</b>	± 5.0		
TGMH MCH	pg	<b>26.4</b>	<b>26.4</b>	± 2.0	<b>29.3</b>	<b>29.3</b>	± 2.0	<b>31.4</b>	<b>31.4</b>	± 2.5		
	fmol	<b>1.64</b>	<b>1.64</b>	± 0.12	<b>1.82</b>	<b>1.82</b>	± 0.12	<b>1.95</b>	<b>1.95</b>	± 0.16		
CCMH MCHC	g/dl	<b>33.6</b>	<b>33.6</b>	± 3.0	<b>34.3</b>	<b>34.3</b>	± 3.0	<b>33.8</b>	<b>33.8</b>	± 3.0		
	g/l	<b>336</b>	<b>336</b>	± 30	<b>343</b>	<b>343</b>	± 30	<b>338</b>	<b>338</b>	± 30		
IDR-SD RDW-SD	mmol/l	<b>20.85</b>	<b>20.85</b>	± 1.86	<b>21.29</b>	<b>21.29</b>	± 1.86	<b>20.96</b>	<b>20.96</b>	± 1.86		
	fl	<b>43.5</b>	<b>43.5</b>	± 8.0	<b>40.5</b>	<b>40.5</b>	± 8.0	<b>40.5</b>	<b>40.5</b>	± 8.0		
IDR-CV RDW-CV	%	<b>14.0</b>	<b>14.0</b>	± 4.0	<b>12.0</b>	<b>12.0</b>	± 4.0	<b>11.5</b>	<b>11.5</b>	± 4.0		
PLA PLT	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>72</b>	<b>72</b>	± 20	<b>235</b>	<b>235</b>	± 30	<b>492</b>	<b>492</b>	± 50		
PLT-Ox	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	N/A	<b>75</b>	± 30	N/A	<b>235</b>	± 40	N/A	<b>470</b>	± 50		
VMP MPV	µm <sup>3</sup> ; fl	<b>8.6</b>	<b>8.6</b>	± 2.0	<b>9.2</b>	<b>9.2</b>	± 2.0	<b>10.0</b>	<b>10.0</b>	± 2.0		
NEU	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>1.28</b>	<b>1.28</b>	± 0.35	<b>4.03</b>	<b>4.03</b>	± 0.90	<b>12.87</b>	<b>12.87</b>	± 1.90		
	%	<b>52.3</b>	<b>52.3</b>	± 10.0	<b>55.2</b>	<b>55.2</b>	± 10.0	<b>72.3</b>	<b>72.3</b>	± 10.0		
LYM	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>0.72</b>	<b>0.72</b>	± 0.33	<b>2.25</b>	<b>2.25</b>	± 0.70	<b>2.60</b>	<b>2.60</b>	± 1.50		
	%	<b>29.2</b>	<b>29.2</b>	± 12.0	<b>30.8</b>	<b>30.8</b>	± 8.0	<b>14.6</b>	<b>14.6</b>	± 8.0		
MON	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>0.17</b>	<b>0.17</b>	± 0.17	<b>0.43</b>	<b>0.43</b>	± 0.39	<b>0.66</b>	<b>0.66</b>	± 0.59		
	%	<b>6.8</b>	<b>6.8</b>	± 6.80	<b>5.9</b>	<b>5.9</b>	± 5.3	<b>3.7</b>	<b>3.7</b>	± 3.3		
EOS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>0.18</b>	<b>0.18</b>	± 0.18	<b>0.28</b>	<b>0.28</b>	± 0.28	<b>0.87</b>	<b>0.87</b>	± 0.87		
	%	<b>7.3</b>	<b>7.3</b>	± 7.30	<b>3.8</b>	<b>3.8</b>	± 3.8	<b>4.9</b>	<b>4.9</b>	± 4.9		
BAS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>0.11</b>	<b>0.11</b>	± 0.11	<b>0.31</b>	<b>0.31</b>	± 0.21	<b>0.80</b>	<b>0.80</b>	± 0.40		
	%	<b>4.4</b>	<b>4.4</b>	± 2.5	<b>4.3</b>	<b>4.3</b>	± 3.0	<b>4.5</b>	<b>4.5</b>	± 2.5		
ERB NRBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>0.42</b>	<b>0.42</b>	± 0.30	<b>1.32</b>	<b>1.32</b>	± 0.45	<b>1.16</b>	<b>1.16</b>	± 0.35		
	%	<b>17.3</b>	<b>17.3</b>	± 8.0	<b>18.1</b>	<b>18.1</b>	± 8.0	<b>6.5</b>	<b>6.5</b>	± 2.0		

BACK / VERSO  
Ref: TEMP-0821 Rev.49