


LOT PX 444
Rev 1

CONTROL

 (Exp.) **2024-01-05**
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500 Since V1.7	H2500 Since V1.7		H1500 Since V1.7	H2500 Since V1.7		H1500 Since V1.7	H2500 Since V1.7		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.57	2.57		± 0.40	7.54		7.54	± 1.00		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.33	± 0.16	4.65	4.65	± 0.20	5.22	5.22	± 0.25	
HB HGB	g/dl	6.2	6.2	± 0.4	13.6	13.6	± 0.5	16.2	16.2	± 0.6	
	g/l	62	62	± 4	136	136	± 5	162	162	± 6	
	mmol/l	3.85	3.85	± 0.25	8.45	8.45	± 0.31	10.06	10.06	± 0.37	
HT HCT	%	18.5	18.5	± 1.5	40.5	40.5	± 2.0	48.8	48.8	± 2.5	
	l/l	0.185	0.185	± 0.015	0.405	0.405	± 0.020	0.488	0.488	± 0.025	
VGM MCV	µm ³ ; fl	79.6	79.6	± 5.0	87.0	87.0	± 5.0	93.5	93.5	± 5.0	
TGMH MCH	pg	26.6	26.6	± 2.0	29.2	29.2	± 2.0	31.0	31.0	± 2.5	
	fmol	1.65	1.65	± 0.12	1.82	1.82	± 0.12	1.93	1.93	± 0.16	
CCMH MCHC	g/dl	33.4	33.4	± 3.0	33.6	33.6	± 3.0	33.2	33.2	± 3.0	
	g/l	334	334	± 30	336	336	± 30	332	332	± 30	
	mmol/l	20.76	20.76	± 1.86	20.88	20.88	± 1.86	20.61	20.61	± 1.86	
IDR-SD RDW-SD	fl	40.3	40.3	± 8.0	37.8	37.8	± 8.0	38.1	38.1	± 8.0	
IDR-CV RDW-CV	%	15.6	15.6	± 4.0	13.4	13.4	± 4.0	12.2	12.2	± 4.0	
PLA PLT	10 ⁹ /mm ³ ; 10 ⁹ /l	73	73	± 20	240	240	± 30	501	501	± 50	
PLT-Ox	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	70	± 30	N/A	227	± 40	N/A	487	± 50	
VMP MPV	µm ³ ; fl	9.7	9.7	± 2.0	9.6	9.6	± 2.0	10.3	10.3	± 2.0	
LPF	%	N/A	2.1	± 2.1	N/A	1.8	± 1.8	N/A	1.9	± 1.9	
NEU	10 ⁹ /mm ³ ; 10 ⁹ /l	1.51	1.51	± 0.35	4.61	4.61	± 0.90	12.73	12.73	± 1.90	
	%	58.9	58.9	± 10.0	61.2	61.2	± 10.0	72.6	72.6	± 10.0	
LYM	10 ⁹ /mm ³ ; 10 ⁹ /l	0.54	0.54	± 0.33	1.64	1.64	± 0.70	1.88	1.88	± 1.50	
	%	21.2	21.2	± 12.0	21.8	21.8	± 8.0	10.7	10.7	± 8.0	
MON	10 ⁹ /mm ³ ; 10 ⁹ /l	0.16	0.16	± 0.16	0.45	0.45	± 0.41	0.63	0.63	± 0.57	
	%	6.3	6.3	± 6.30	6.0	6.0	± 5.4	3.6	3.6	± 3.2	
EOS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.16	0.16	± 0.16	0.26	0.26	± 0.26	0.79	0.79	± 0.79	
	%	6.1	6.1	± 6.10	3.4	3.4	± 3.4	4.5	4.5	± 4.5	
BAS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.12	0.12	± 0.12	0.34	0.34	± 0.23	0.81	0.81	± 0.41	
	%	4.5	4.5	± 2.5	4.5	4.5	± 3.0	4.6	4.6	± 2.5	
IMG	10 ⁹ /mm ³ ; 10 ⁹ /l	0.08	0.08	± 0.08	0.23	0.23	± 0.21	0.70	0.70	± 0.63	
	%	3.0	3.0	± 3.00	3.1	3.1	± 2.8	4.0	4.0	± 3.6	
ERB NRBC	10 ⁹ /mm ³ ; 10 ⁹ /l	0.23	0.23	± 0.23	0.71	0.71	± 0.45	0.81	0.81	± 0.35	
	%	9.1	9.1	± 8.0	9.4	9.4	± 8.0	4.6	4.6	± 2.0	
TNC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.80	2.80	± 0.40	8.25	8.25	± 1.00	18.34	18.34	± 2.20	

Ref: TEMP-0821 Rev.49