

LOT PX 445
Rev 1

CONTROL

 (Exp.) **2024-03-05**
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio									
		CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
		YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
		H1500 HELO 2.0	H2500 HELO 2.0		H1500 HELO 2.0	H2500 HELO 2.0		H1500 HELO 2.0	H2500 HELO 2.0		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.62	2.62		± 0.40	7.70		7.70	± 1.00		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40	± 0.16	4.67	4.67	± 0.20	5.15	5.15	± 0.25	
HB HGB	g/dl	6.2	6.2	± 0.4	13.7	13.7	± 0.5	16.2	16.2	± 0.6	
	g/l	62	62	± 4	137	137	± 5	162	162	± 6	
	mmol/l	3.85	3.85	± 0.25	8.51	8.51	± 0.31	10.06	10.06	± 0.37	
HT HCT	%	18.5	18.5	± 1.5	40.5	40.5	± 2.0	48.5	48.5	± 2.5	
	l/l	0.185	0.185	± 0.015	0.405	0.405	± 0.020	0.485	0.485	± 0.025	
VGM MCV	µm ³ ; fl	77.0	77.0	± 5.0	86.9	86.9	± 5.0	94.1	94.1	± 5.0	
TGMH MCH	pg	25.8	25.8	± 2.0	29.3	29.3	± 2.0	31.5	31.5	± 2.5	
	fmol	1.60	1.60	± 0.12	1.82	1.82	± 0.12	1.95	1.95	± 0.16	
CCMH MCHC	g/dl	33.5	33.5	± 3.0	33.8	33.8	± 3.0	33.4	33.4	± 3.0	
	g/l	335	335	± 30	338	338	± 30	334	334	± 30	
	mmol/l	20.83	20.83	± 1.86	20.96	20.96	± 1.86	20.76	20.76	± 1.86	
IDR-SD RDW-SD	fl	43.0	43.0	± 8.0	40.2	40.2	± 8.0	41.0	41.0	± 8.0	
IDR-CV RDW-CV	%	15.8	15.8	± 4.0	13.0	13.0	± 4.0	11.6	11.6	± 4.0	
PLA PLT	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72	± 20	243	243	± 30	486	486	± 50	
PLT-Ox	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	72	± 30	N/A	237	± 40	N/A	476	± 50	
VMP MPV	µm ³ ; fl	10.1	10.1	± 2.0	9.8	9.8	± 2.0	10.2	10.2	± 2.0	
LPF	%	N/A	2.4	± 2.4	N/A	1.7	± 1.7	N/A	1.9	± 1.9	
NEU	10 ⁹ /mm ³ ; 10 ⁹ /l	1.49	1.49	± 0.35	4.63	4.63	± 0.90	12.84	12.84	± 1.90	
	%	56.7	56.7	± 10.0	60.1	60.1	± 10.0	71.5	71.5	± 10.0	
LYM	10 ⁹ /mm ³ ; 10 ⁹ /l	0.59	0.59	± 0.33	1.81	1.81	± 0.70	1.99	1.99	± 1.50	
	%	22.4	22.4	± 12.0	23.5	23.5	± 8.0	11.1	11.1	± 8.0	
MON	10 ⁹ /mm ³ ; 10 ⁹ /l	0.21	0.21	± 0.21	0.45	0.45	± 0.41	0.84	0.84	± 0.76	
	%	8.0	8.0	± 8.00	5.9	5.9	± 5.3	4.7	4.7	± 4.2	
EOS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.15	0.15	± 0.15	0.25	0.25	± 0.25	0.74	0.74	± 0.74	
	%	5.7	5.7	± 5.70	3.2	3.2	± 3.2	4.1	4.1	± 4.1	
BAS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.12	0.12	± 0.12	0.33	0.33	± 0.22	0.84	0.84	± 0.42	
	%	4.4	4.4	± 2.5	4.3	4.3	± 3.0	4.7	4.7	± 2.5	
IMG	10 ⁹ /mm ³ ; 10 ⁹ /l	0.07	0.07	± 0.07	0.23	0.23	± 0.21	0.70	0.70	± 0.63	
	%	2.8	2.8	± 2.80	3.0	3.0	± 2.7	3.9	3.9	± 3.5	
ERB NRBC	10 ⁹ /mm ³ ; 10 ⁹ /l	0.25	0.25	± 0.25	0.78	0.78	± 0.45	0.86	0.86	± 0.35	
	%	9.6	9.6	± 8.0	10.1	10.1	± 8.0	4.8	4.8	± 2.0	
TNC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.87	2.87	± 0.40	8.48	8.48	± 1.00	18.82	18.82	± 2.20	

Ref: TEMP-0821 Rev.50