

# ABX Minotrol Retic



**LOT** RX 439  
Rev 1

**CONTROL**

(Exp.) 2023-03-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
Méthode Manuelle / Manual RETIC	%	<b>1.4</b>			± 0.7	<b>5.9</b>			± 2.4	<b>11.0</b>			± 3.3
RETIC	%	<b>1.61</b>			± 0.81	<b>6.06</b>			± 1.52	<b>11.81</b>			± 2.36
RETIC	10 <sup>6</sup> /mm <sup>3</sup>	<b>0.056</b>			± 0.028	<b>0.211</b>			± 0.053	<b>0.413</b>			± 0.083
	10 <sup>9</sup> /l	<b>56</b>			± 28	<b>211</b>			± 53	<b>413</b>			± 83
MFI MRI *	%	<b>30.5</b>			± 8.0	<b>42.9</b>			± 10.0	<b>46.4</b>			± 12.0
	-	<b>0.305</b>			± 0.080	<b>0.429</b>			± 0.100	<b>0.464</b>			± 0.120
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>3.50</b>			± 0.30	<b>3.49</b>			± 0.30	<b>3.50</b>			± 0.30
PIC RANGE	-	<b>13</b>			± 8	<b>13</b>			± 8	<b>13</b>			± 8

1300102875  
Ref: TEMP-0382 Rev.42

\* For research use only in Canada and USA