

# ABX Minotrol Retic



**LOT** RX 444  
Rev 1

**CONTROL**

(Exp.) 2024-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
		H2500 V1.0 to V1.4				H2500 V1.0 to V1.4				H2500 V1.0 to V1.4			
Méthode Manuelle / Manual													
RET	%	<b>1.6</b>			± 0.8	<b>4.9</b>			± 2.0	<b>10.5</b>			± 3.2
RET	%	<b>1.64</b>			± 0.82	<b>5.18</b>			± 1.30	<b>10.86</b>			± 2.17
RET	10 <sup>6</sup> /mm <sup>3</sup>	<b>0.057</b>			± 0.029	<b>0.180</b>			± 0.045	<b>0.372</b>			± 0.074
	10 <sup>9</sup> /l	<b>57</b>			± 29	<b>180</b>			± 45	<b>372</b>			± 74
MFI MRI *	%	<b>32.6</b>			± 8.0	<b>44.7</b>			± 10.0	<b>47.9</b>			± 12.0
	-	<b>0.326</b>			± 0.080	<b>0.447</b>			± 0.100	<b>0.479</b>			± 0.120
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>3.45</b>			± 0.30	<b>3.47</b>			± 0.30	<b>3.43</b>			± 0.30
PIC RANGE	-	<b>19</b>			± 8	<b>19</b>			± 8	<b>19</b>			± 8

Ref: TEMP-0382 Rev.44 FRONT / RECTO 1300116193

\* For research use only in Canada and USA