

ABX Minotrol Retic



LOT RX 443

CONTROL

(Exp.) **2023-11-05**
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
		H2500 Since V1.7				H2500 Since V1.7				H2500 Since V1.7			
Méthode Manuelle / Manual													
RET	%	1.4			± 0.7	5.5			± 2.3	11.8			± 3.3
RET	%	1.81			± 0.91	5.41			± 1.35	10.44			± 2.09
RET	10 ⁶ /mm ³	0.062			± 0.031	0.183			± 0.046	0.350			± 0.070
	10 ⁹ /l	62			± 31	183			± 46	350			± 70
CRC	%	1.11			± 0.56	3.41			± 1.02	6.45			± 1.61
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	3.42			± 0.30	3.39			± 0.30	3.35			± 0.30
IRF	--	0.41			± 0.25	0.53			± 0.32	0.58			± 0.35
RHCC	pg	21.5			± 4.3	15.7			± 3.1	12.3			± 3.1

Ref: TEMP-0382 Rev.44 BACK / VERSO 1300116193