

LOT BFI 2444
Rev 1

CONTROL

 (Exp.) 2024-01-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | BFTROL | | | | | | | | | |
|--------------------------|--------------------|------------------|------------------|------------------|------------------|-------------------------|------------------|------------------|------------------|------------------|-------------------------|
| | | CONTROL | | | | TOLERANCES TOLERANCE | CONTROL | | | | TOLERANCES TOLERANCE |
| | | RANGE | | RANGE | | | RANGE | | RANGE | | |
| | | YUMIZEN H1500 | YUMIZEN H2500 | YUMIZEN H1500 | YUMIZEN H2500 | | YUMIZEN H1500 | YUMIZEN H2500 | YUMIZEN H1500 | YUMIZEN H2500 | |
| | | Since V1.7 | | | Since V1.7 | | | | | | |
| BFWBC | 10 ⁶ /l | 237 | 237 | | ± 95 | 908 | 908 | | | ± 227 | |
| BFRBC | 10 ⁶ /l | 77691 | 77691 | | ± 19423 | 508674 | 508674 | | | ± 127169 | |
| BFMN | 10 ⁶ /l | 68 | 68 | | ± 24 | 102 | 102 | | | ± 41 | |
| | % | 28 | 28 | | ± 10 | 11 | 11 | | | ± 4 | |
| BFPN | 10 ⁶ /l | 169 | 169 | | ± 59 | 806 | 806 | | | ± 202 | |
| | % | 72 | 72 | | ± 25 | 89 | 89 | | | ± 22 | |

Ref: TEMP-5502 Rev.06