



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 3 0 8 1 6 4 6 4 1 2 0 8 2 5 D



D 5 1 0 6 1 4 3 0 7 7 1 2 8 0 7 9 D



D 6 1 1 9 1 0 3 1 4 0 0 8 6 1 4 3 D



D 7 0 7 5 1 1 3 0 8 4 1 1 3 0 8 7 D



D 8 1 4 4 0 7 5 1 1 3 3 4 4 4 6 5 D



D 9 2 9 1 3 9 4 1 8 4 2 7 6 2 2 3 D



D 0 3 3 4 3 7 0 D

Reference No.:		1300036412	
Lot No:		930816 (rev. 3)	
Expiry date:		2025-08-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,6-14,3
	Prothrombin %	%	77-128
	Prothrombin INR	INR	0,79-1,19
PT Liq	Prothrombin Time	sec	10,3-14,0
	Prothrombin %	%	86-143
	Prothrombin INR	INR	0,75-1,13
PT R	Prothrombin Time	sec	8,4-11,3
	Prothrombin %	%	87-144
	Prothrombin INR	INR	0,75-1,13
APTT	Partial Thromboplastin Time	sec	34,4-46,5
APTT Liq	Partial Thromboplastin Time	sec	29,1-39,4
FIB Cl	Fibrinogen	g/L	1,84-2,76
TT	Thrombin Time	sec	22,3-33,4
AT	Antithrombin	%	90-122

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 3 0 8 1 6 5 6 4 1 2 0 8 2 5 D



D 5 1 5 3 2 3 0 0 3 1 0 5 1 1 3 2 D



D 6 1 9 8 1 7 7 2 6 6 0 2 8 0 4 7 D



D 7 1 5 3 2 2 9 1 2 3 1 8 4 0 3 8 D



D 8 0 6 3 1 1 8 1 7 7 6 6 3 8 9 8 D



D 9 6 1 3 8 3 0 0 6 9 1 6 0 0 0 0 D



D 0 9 9 9 3 9 3 D

Reference No.:		1300036412	
Lot No:		930816 (rev. 3)	
Expiry date:		2025-08-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	15,3-23,0
	Prothrombin %	%	31-51
	Prothrombin INR	INR	1,32-1,98
PT Liq	Prothrombin Time	sec	17,7-26,6
	Prothrombin %	%	28-47
	Prothrombin INR	INR	1,53-2,29
PT R	Prothrombin Time	sec	12,3-18,4
	Prothrombin %	%	38-63
	Prothrombin INR	INR	1,18-1,77
APTT	Partial Thromboplastin Time	sec	66,3-89,8
APTT Liq	Partial Thromboplastin Time	sec	61,3-83,0
FIB Cl	Fibrinogen	g/L	0,69-1,60
TT	Thrombin Time	sec	N/A
AT	Antithrombin	%	34-57
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			