

**LOT** MX 446 L  
Rev 3

**CONTROL** L

 (Exp.)

**2024-06-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST				60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0				2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.26				2.23					± 0.17
HB HGB	g/dl	6.1				6.1					± 0.5
	g/l	61				61					± 5
	mmol/l	3.79				3.79					± 0.31
HT HCT	%	16.5				16.3					± 2.0
	l/l	0.165				0.163					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	73				73					± 4
TGMH MCH	pg	27.0				27.4					± 2.0
	fmol	1.68				1.70					± 0.12
CCMH MCHC	g/dl	37.0				37.5					± 3.0
	g/l	370				375					± 30
	mmol/l	22.96				23.27					± 1.86
IDR RDW	%	15.0				14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75				73					± 22
VPM MPV	µm <sup>3</sup> ; fl	8.6				7.5					± 2.0
LYMPHO	#	1.27				1.23					± 0.40
	%	63.3				61.5					± 8.0
MONO	#	0.20				0.23					± 0.20
	%	9.6				11.7					± 6.0
GRANULO	#	0.54				0.54					± 0.40
	%	27.1				26.8					± 7.0

**LOT** MX 446 N  
Rev 3

**CONTROL** N

(Exp.)

**2024-06-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST				60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.7				7.7					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.54				4.55					± 0.26
HB HGB	g/dl	13.6				13.6					± 0.7
	g/l	136				136					± 7
	mmol/l	8.45				8.45					± 0.43
HT HCT	%	37.7				37.8					± 2.7
	l/l	0.377				0.378					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	83				83					± 4
TGMH MCH	pg	30.0				29.9					± 2.0
	fmol	1.86				1.86					± 0.12
CCMH MCHC	g/dl	36.1				36.0					± 3.0
	g/l	361				360					± 30
	mmol/l	22.41				22.36					± 1.86
IDR RDW	%	14.5				13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	275				260					± 40
VPM MPV	µm <sup>3</sup> ; fl	7.7				7.5					± 2.0
LYMPHO	#	2.49				2.49					± 0.40
	%	32.4				32.4					± 5.0
MONO	#	1.18				1.18					± 1.06
	%	15.3				15.3					± 13.8
GRANULO	#	4.03				4.03					± 0.77
	%	52.3				52.3					± 10.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104144

**LOT** MX 446 H  
Rev 3

**CONTROL** H

 (Exp.)

**2024-06-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST				60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.5</b>				<b>20.5</b>					± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.65</b>				<b>5.68</b>					± 0.32
HB HGB	g/dl	<b>18.1</b>				<b>18.1</b>					± 0.9
	g/l	<b>181</b>				<b>181</b>					± 9
	mmol/l	<b>11.24</b>				<b>11.24</b>					± 0.56
HT HCT	%	<b>50.9</b>				<b>51.1</b>					± 3.4
	l/l	<b>0.509</b>				<b>0.511</b>					± 0.034
VGM MCV	µm <sup>3</sup> ; fl	<b>90</b>				<b>90</b>					± 4
TGMH MCH	pg	<b>32.0</b>				<b>31.9</b>					± 2.0
	fmol	<b>1.99</b>				<b>1.98</b>					± 0.12
CCMH MCHC	g/dl	<b>35.6</b>				<b>35.4</b>					± 3.0
	g/l	<b>356</b>				<b>354</b>					± 30
	mmol/l	<b>22.10</b>				<b>21.99</b>					± 1.86
IDR RDW	%	<b>14.5</b>				<b>13.0</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>550</b>				<b>530</b>					± 65
VPM MPV	µm <sup>3</sup> ; fl	<b>7.7</b>				<b>7.5</b>					± 2.0
LYMPHO	#	<b>3.28</b>				<b>3.28</b>					± 1.00
	%	<b>16.0</b>				<b>16.0</b>					± 5.0
MONO	#	<b>1.03</b>				<b>1.03</b>					± 0.80
	%	<b>5.0</b>				<b>5.0</b>					± 4.0
GRANULO	#	<b>16.20</b>				<b>16.20</b>					± 1.40
	%	<b>79.0</b>				<b>79.0</b>					± 7.0