

LOT

MX 446 L
Rev 3

CONTROL L

 (Exp.)

2024-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60				60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0				2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.24				2.23					± 0.17
HB HGB	g/dl	5.9				5.7					± 0.5
	g/l	59				57					± 5
	mmol/l	3.66				3.52					± 0.31
HT HCT	%	16.4				16.5					± 2.0
	l/l	0.164				0.165					± 0.020
VGM MCV	µm ³ ; fl	73				74					± 4
TGMH MCH	pg	26.3				25.4					± 2.0
	fmol	1.64				1.58					± 0.12
CCMH MCHC	g/dl	36.1				34.4					± 3.0
	g/l	361				344					± 30
	mmol/l	22.41				21.35					± 1.86
IDR RDW	%	15.0				15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70				73					± 22
VPM MPV	µm ³ ; fl	8.8				7.5					± 2.0
LYMPHO	#	1.27				1.28					± 0.40
	%	63.3				64.0					± 8.0
MONO	#	0.20				0.20					± 0.20
	%	9.6				9.0					± 6.0
GRANULO	#	0.54				0.54					± 0.40
	%	27.1				27.0					± 7.0

LOT

MX 446 N
Rev 3

CONTROL N

 (Exp.)

2024-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60				60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7				7.5					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.52				4.51					± 0.26
HB HGB	g/dl	13.6				13.3					± 0.7
	g/l	136				133					± 7
	mmol/l	8.45				8.26					± 0.43
HT HCT	%	37.5				37.9					± 2.7
	l/l	0.375				0.379					± 0.027
VGM MCV	µm ³ ; fl	83				84					± 4
TGMH MCH	pg	30.1				29.5					± 2.0
	fmol	1.87				1.83					± 0.12
CCMH MCHC	g/dl	36.3				35.1					± 3.0
	g/l	363				351					± 30
	mmol/l	22.51				21.80					± 1.86
IDR RDW	%	14.5				14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	259				260					± 40
VPM MPV	µm ³ ; fl	8.0				7.5					± 2.0
LYMPHO	#	2.45				2.45					± 0.40
	%	32.7				32.7					± 5.0
MONO	#	1.15				1.15					± 1.04
	%	15.3				15.3					± 13.8
GRANULO	#	3.90				3.90					± 0.75
	%	52.0				52.0					± 10.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

LOT

MX 446 H
Rev 3

CONTROL H

 (Exp.)

2024-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60				60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.3				19.9					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.62				5.61					± 0.32
HB HGB	g/dl	18.4				17.9					± 0.9
	g/l	184				179					± 9
	mmol/l	11.43				11.13					± 0.56
HT HCT	%	50.0				50.5					± 3.4
	l/l	0.500				0.505					± 0.034
VGM MCV	µm ³ ; fl	89				90					± 4
TGMH MCH	pg	32.7				31.9					± 2.0
	fmol	2.03				1.98					± 0.12
CCMH MCHC	g/dl	36.8				35.5					± 3.0
	g/l	368				355					± 30
	mmol/l	22.84				22.04					± 1.86
IDR RDW	%	14.0				14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	515				530					± 65
VPM MPV	µm ³ ; fl	8.0				7.5					± 2.0
LYMPHO	#	3.45				3.38					± 1.00
	%	17.0				17.0					± 5.0
MONO	#	0.91				0.90					± 0.80
	%	4.5				4.5					± 4.0
GRANULO	#	15.94				15.62					± 1.40
	%	78.5				78.5					± 7.0