

ABX Erytrol



LOT

NX 446

CONTROL



(Exp.) 2024-05-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA			PENTRA	PENTRA			PENTRA	PENTRA		
		DX120	DX NEXUS	DX120	DX NEXUS	DX120	DX NEXUS						
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	N/A	7.4		± 0.8	N/A	18.2		± 1.5	N/A	2.7		± 0.4
GBC CWBC	10 ³ /mm ³ ; 10 ⁹ /l	N/A	7.2		± 0.8	N/A	16.7		± 1.4	N/A	1.5		± 0.4
ERB	%	N/A	3.2		± 1.7	N/A	9.0		± 3.3	N/A	82.8		± 50.0
ERB	#	N/A	0.230		± 0.100	N/A	1.503		± 0.700	N/A	1.242		± 0.500

1219930076-C

Ref: TEMP-0822 Rev.36

GBC / CWBC

* les valeurs sont données à titre d'indication et doivent être utilisées comme RuO (Utilisation en Recherche Seulement) Indicative expected values, must be used as RuO (Research used only)