


LOT PX 446
Rev 1


CONTROL

 (Exp.) 2024-05-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.1	3.1			± 0.4	8.7	8.7			± 1.0	18.7	18.7			± 2.2			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.28	2.28			± 0.16	4.50	4.50			± 0.20	5.17	5.17			± 0.25			
HB HGB	g/dl	5.9	5.9			± 0.4	13.2	13.2			± 0.5	16.7	16.7			± 0.6			
	g/l	59	59			± 4	132	132			± 5	167	167			± 6			
	mmol/l	3.66	3.66			± 0.25	8.20	8.20			± 0.31	10.37	10.37			± 0.37			
HT HCT	%	17.8	17.8			± 1.5	38.3	38.3			± 2.0	47.6	47.6			± 2.5			
	l/l	0.178	0.178			± 0.015	0.383	0.383			± 0.020	0.476	0.476			± 0.025			
VGM MCV	µm ³ fl	78	78			± 5	85	85			± 5	92	92			± 5			
TGMH MCH	pg	25.9	25.9			± 2.0	29.3	29.3			± 2.0	32.3	32.3			± 2.5			
	fmol	1.61	1.61			± 0.12	1.82	1.82			± 0.12	2.01	2.01			± 0.16			
CCMH MCHC	g/dl	33.2	33.2			± 3.0	34.5	34.5			± 3.0	35.1	35.1			± 3.0			
	g/l	332	332			± 30	345	345			± 30	351	351			± 30			
	mmol/l	20.60	20.60			± 1.86	21.43	21.43			± 1.86	21.80	21.80			± 1.86			
IDR RDW	%	14.5	14.5			± 4.0	12.5	12.5			± 4.0	12.5	12.5			± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	62	62			± 20	225	225			± 30	440	440			± 50			
VPM MPV	µm ³ fl	9.5	9.5			± 2.0	9.2	9.2			± 2.0	10.1	10.1			± 2.0			
NEUT	#	1.56	1.56			± 0.35	4.79	4.79			± 0.90	12.60	12.60			± 1.90			
	%	50.2	50.2			± 10.0	55.0	55.0			± 10.0	67.4	67.4			± 10.0			
LYMPHO	#	1.11	1.11			± 0.33	3.06	3.06			± 0.70	3.98	3.98			± 1.50			
	%	35.9	35.9			± 12.0	35.2	35.2			± 8.0	21.3	21.3			± 8.0			
MONO	#	0.16	0.16			± 0.16	0.30	0.30			± 0.30	0.52	0.52			± 0.52			
	%	5.2	5.2			± 5.2	3.5	3.5			± 3.5	2.8	2.8			± 2.8			
EOS	#	0.18	0.18			± 0.18	0.29	0.29			± 0.29	0.88	0.88			± 0.88			
	%	5.9	5.9			± 5.9	3.3	3.3			± 3.3	4.7	4.7			± 4.7			
BASO	#	0.09	0.09			± 0.09	0.26	0.26			± 0.26	0.71	0.71			± 0.71			
	%	2.8	2.8			± 2.8	3.0	3.0			± 3.0	3.8	3.8			± 3.8			

LOT PX 446
Rev 1

CONTROL

 (Exp.) 2024-05-05
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCRP		60 60C+ ES60		MS60	MSCRP		60 60C+ ES60	MS60		MSCRP		60 60C+ ES60	MS60	MSCRP	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0	3.0		± 0.4	8.6	8.6	8.6		± 1.0	18.8	18.6	18.5		± 2.2			
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.35	2.33	2.32		± 0.16	4.52	4.51	4.48		± 0.20	5.19	5.14	5.15		± 0.25			
	g/dl	6.0	6.0	6.0		± 0.4	13.3	13.3	13.2		± 0.5	16.7	16.6	16.6		± 0.6			
HB HGB	g/l	60	60	60		± 4	133	133	132		± 5	167	166	166		± 6			
	mmol/l	3.73	3.73	3.73		± 0.25	8.26	8.26	8.20		± 0.31	10.37	10.31	10.31		± 0.37			
HT HCT	%	17.9	17.5	17.5		± 1.5	38.4	37.4	37.6		± 2.0	48.3	46.8	46.9		± 2.5			
	l/l	0.179	0.175	0.175		± 0.015	0.384	0.374	0.376		± 0.020	0.483	0.468	0.469		± 0.025			
VGM MCV	µm ³ ; fl	76	75	75.5		± 5	85	83	84.0		± 5	93	91	91.0		± 5			
TGMH MCH	pg	25.5	25.8	25.9		± 2.0	29.4	29.5	29.5		± 2.0	32.2	32.3	32.2		± 2.5			
	fmol	1.59	1.60	1.61		± 0.12	1.83	1.83	1.83		± 0.12	2.00	2.01	2.00		± 0.16			
	g/dl	33.6	34.3	34.3		± 3.0	34.6	35.5	35.1		± 3.0	34.6	35.5	35.4		± 3.0			
CCMH MCHC	g/l	336	343	343		± 30	346	355	351		± 30	346	355	354		± 30			
	mmol/l	20.86	21.32	21.27		± 1.86	21.50	22.06	21.78		± 1.86	21.49	22.04	22.00		± 1.86			
IDR RDW	%	15.0	15.0	13.0		± 4.0	12.5	12.5	11.5		± 4.0	12.0	12.0	11.0		± 4.0			
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	64	66		± 20	230	226	221		± 30	439	450	428		± 50			
VPM MPV	µm ³ ; fl	9.2	9.2	9.0		± 2.0	8.6	8.5	8.6		± 2.0	9.6	9.6	9.4		± 2.0			
NEUT	#	1.51	1.43	1.52		± 0.35	4.71	4.64	4.73		± 0.90	12.52	12.39	12.67		± 1.90			
	%	50.3	47.8	50.5		± 10.0	54.8	53.9	55.0		± 10.0	66.6	66.6	68.5		± 10.0			
LYMPHO	#	1.09	1.17	1.08		± 0.33	3.04	3.18	3.10		± 0.70	4.04	4.19	3.89		± 1.50			
	%	36.3	39.0	36.0		± 12.0	35.3	37.0	36.0		± 8.0	21.5	22.5	21.0		± 8.0			
MONO	#	0.14	0.12	0.15		± 0.12	0.29	0.24	0.26		± 0.24	0.53	0.43	0.37		± 0.37			
	%	4.8	4.1	5.0		± 4.1	3.4	2.8	3.0		± 2.8	2.8	2.3	2.0		± 2.0			
EOS	#	0.17	0.19	0.17		± 0.17	0.30	0.28	0.26		± 0.26	0.98	0.89	0.83		± 0.83			
	%	5.8	6.3	5.5		± 5.5	3.5	3.3	3.0		± 3.0	5.2	4.8	4.5		± 4.5			
BASO	#	0.08	0.08	0.09		± 0.08	0.26	0.26	0.26		± 0.26	0.73	0.71	0.74		± 0.71			
	%	2.8	2.8	3.0		± 2.8	3.0	3.0	3.0		± 3.0	3.9	3.8	4.0		± 3.8			

Ref: TEMP-0821 Rev.50 FRONT / RECTO