

LOT

MX 446 L
Rev 3

CONTROL L

 (Exp.)

2024-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST				60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0				2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.26				2.23					± 0.17
HB HGB	g/dl	6.1				6.1					± 0.5
	g/l	61				61					± 5
	mmol/l	3.79				3.79					± 0.31
HT HCT	%	16.5				16.3					± 2.0
	l/l	0.165				0.163					± 0.020
VGM MCV	µm ³ ; fl	73				73					± 4
TGMH MCH	pg	27.0				27.4					± 2.0
	fmol	1.68				1.70					± 0.12
CCMH MCHC	g/dl	37.0				37.5					± 3.0
	g/l	370				375					± 30
	mmol/l	22.96				23.27					± 1.86
IDR RDW	%	15.0				14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75				73					± 22
VPM MPV	µm ³ ; fl	8.6				7.5					± 2.0
LYMPHO	#	1.27				1.23					± 0.40
	%	63.3				61.5					± 8.0
MONO	#	0.20				0.23					± 0.20
	%	9.6				11.7					± 6.0
GRANULO	#	0.54				0.54					± 0.40
	%	27.1				26.8					± 7.0

LOT		MX 446 N	CONTROL N				(Exp.)				2024-06-05 (YYYY-MM-DD)	
		Rev 3	ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		ES60 Care ST				60 8P - 60 18P ADVIA 60						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7				7.7					± 0.8	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.54				4.55					± 0.26	
HB HGB	g/dl	13.6				13.6					± 0.7	
	g/l	136				136					± 7	
	mmol/l	8.45				8.45					± 0.43	
HT HCT	%	37.7				37.8					± 2.7	
	l/l	0.377				0.378					± 0.027	
VGM MCV	µm ³ ; fl	83				83					± 4	
TGMH MCH	pg	30.0				29.9					± 2.0	
	fmol	1.86				1.86					± 0.12	
CCMH MCHC	g/dl	36.1				36.0					± 3.0	
	g/l	361				360					± 30	
	mmol/l	22.41				22.36					± 1.86	
IDR RDW	%	14.5				13.5					± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	275				260					± 40	
VPM MPV	µm ³ ; fl	7.7				7.5					± 2.0	
LYMPHO	#	2.49				2.49					± 0.40	
	%	32.4				32.4					± 5.0	
MONO	#	1.18				1.18					± 1.06	
	%	15.3				15.3					± 13.8	
GRANULO	#	4.03				4.03					± 0.77	
	%	52.3				52.3					± 10.0	

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104144

LOT

MX 446 H
 Rev 3

CONTROL H

(Exp.)

2024-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST				60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.5				20.5					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.65				5.68					± 0.32
HB HGB	g/dl	18.1				18.1					± 0.9
	g/l	181				181					± 9
	mmol/l	11.24				11.24					± 0.56
HT HCT	%	50.9				51.1					± 3.4
	l/l	0.509				0.511					± 0.034
VGM MCV	µm ³ ; fl	90				90					± 4
TGMH MCH	pg	32.0				31.9					± 2.0
	fmol	1.99				1.98					± 0.12
CCMH MCHC	g/dl	35.6				35.4					± 3.0
	g/l	356				354					± 30
	mmol/l	22.10				21.99					± 1.86
IDR RDW	%	14.5				13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	550				530					± 65
VPM MPV	µm ³ ; fl	7.7				7.5					± 2.0
LYMPHO	#	3.28				3.28					± 1.00
	%	16.0				16.0					± 5.0
MONO	#	1.03				1.03					± 0.80
	%	5.0				5.0					± 4.0
GRANULO	#	16.20				16.20					± 1.40
	%	79.0				79.0					± 7.0