


LOT		MX 446 L		CONTROL		L		 (Exp.)		2024-06-05 (YYYY-MM-DD)			
		Rev 3		ABX Lysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS										TOLERANCES TOLERANCE	
		EMI CRP											
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l											± 0.4	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l											± 0.15	
HB HGB	g/dl											± 0.4	
	g/l											± 4	
	mmol/l											± 0.25	
HT HCT	%											± 2.0	
	l/l											± 0.020	
VGM MCV	µm ³ ; fl											± 4	
TGMH MCH	pg											± 2.0	
	fmol											± 0.12	
CCMH MCHC	g/dl											± 3.0	
	g/l											± 30	
	mmol/l											± 1.86	
IDR RDW	%											± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l											± 20	
VPM MPV	µm ³ ; fl											± 2.0	
LYMPHO	#											± 0.40	
	%											± 8.0	
MONO	#											± 0.20	
	%											± 6.0	
GRANULO	#											± 0.40	
	%											± 7.0	

LOT

MX 446 N
Rev 3

CONTROL

N

 (Exp.)

2024-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
		MICROS								
		EMI CRP								
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			7.9						± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			4.50						± 0.18
HB HGB	g/dl			13.4						± 0.5
	g/l			134						± 5
	mmol/l			8.32						± 0.31
HT HCT	%			36.9						± 2.5
	l/l			0.369						± 0.025
VGM MCV	µm ³ ; fl			82						± 4
TGMH MCH	pg			29.8						± 2.0
	fmol			1.85						± 0.12
CCMH MCHC	g/dl			36.3						± 3.0
	g/l			363						± 30
	mmol/l			22.55						± 1.86
IDR RDW	%			13.5						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			259						± 40
VPM MPV	µm ³ ; fl			7.8						± 2.0
LYMPHO	#			2.53						± 0.40
	%			32.0						± 5.0
MONO	#			1.58						± 1.42
	%			20.0						± 18.0
GRANULO	#			3.79						± 0.79
	%			48.0						± 10.0

Ret: TEMP-0830 Rev.42 BACK / VERSO 1300104144

LOT

MX 446 H
Rev 3

CONTROL

H

 (Exp.)

2024-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			20.6						± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			5.54						± 0.20
HB HGB	g/dl			17.7						± 0.6
	g/l			177						± 6
	mmol/l			10.99						± 0.37
HT HCT	%			49.3						± 3.0
	l/l			0.493						± 0.030
VGM MCV	µm ³ ; fl			89						± 4
TGMH MCH	pg			31.9						± 2.0
	fmol			1.98						± 0.12
CCMH MCHC	g/dl			35.9						± 3.0
	g/l			359						± 30
	mmol/l			22.29						± 1.86
IDR RDW	%			13.0						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			519						± 45
VPM MPV	µm ³ ; fl			7.9						± 2.0
LYMPHO	#			3.26						± 1.00
	%			15.8						± 5.0
MONO	#			1.44						± 0.80
	%			7.0						± 4.0
GRANULO	#			15.90						± 1.40
	%			77.2						± 7.0